

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012523\
 Data File : VD075203.D
 Acq On : 25 Jan 2023 16:59
 Operator : KP/SY
 Sample : 01293-01
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PL-3-012523

Quant Time: Jan 26 03:50:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 19 01:32:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	276	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	443	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	212	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.52
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.216	65	115	40.833	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	81.660%
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	10.269	98	804	76.281	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	152.560%#
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	25 - 144	Recovery	=	0.000%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012523\
Data File : VD075203.D
Acq On : 25 Jan 2023 16:59
Operator : KP/SY
Sample : 01293-01
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PL-3-012523

Quant Time: Jan 26 03:50:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 19 01:32:27 2023
Response via : Initial Calibration

