

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012621\
 Data File : VD068187.D
 Acq On : 26 Jan 2021 13:13
 Operator : VA/SY
 Sample : M1235-04
 Misc : 5.91G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FK-SF-SAMPLE-2

Quant Time: Jan 27 00:11:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	541485	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	963329	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	822283	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	355084	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	276907	53.46	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.92%
35) Dibromofluoromethane	7.920	113	250167	46.97	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.94%
50) Toluene-d8	10.344	98	1153963	54.17	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.34%
62) 4-Bromofluorobenzene	12.632	95	364197	47.47	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.94%
Target Compounds						Qvalue
16) Acetone	4.161	43	154590	167.14	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012621\
Data File : VD068187.D
Acq On : 26 Jan 2021 13:13
Operator : VA/SY
Sample : M1235-04
Misc : 5.91G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FK-SF-SAMPLE-2

Quant Time: Jan 27 00:11:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

