

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012621\
 Data File : VD068193.D
 Acq On : 26 Jan 2021 18:23
 Operator : VA/SY
 Sample : M1236-05
 Misc : 6.52G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 2922

Quant Time: Jan 27 00:13:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	521811	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	920803	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	808422	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.579	152	346062	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	291346	58.37	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.74%
35) Dibromofluoromethane	7.920	113	278545	54.72	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.44%
50) Toluene-d8	10.344	98	1142916	56.13	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.26%
62) 4-Bromofluorobenzene	12.638	95	363446	49.56	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.12%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012621\
Data File : VD068193.D
Acq On : 26 Jan 2021 18:23
Operator : VA/SY
Sample : M1236-05
Misc : 6.52G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
2922

Quant Time: Jan 27 00:13:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

