



Tert butylalcohol



Response = $3.139 \times 10^{-2} \times \text{Amt} + 3.129 \times 10^{-2}$
 Coef of Det (r^2) = 0.994408 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Calibration Table Last Updated: Thu Jan 27 06:04:38 2022

Methylene Chloride

Response Ratio



$$\text{Response} = 5.842\text{e-}001 * \text{Amt} + 4.723\text{e-}002$$


Coef of Det (r^2) = 0.999194 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M

Calibration Table Last Updated: Thu Jan 27 06:04:38 2022

Naphthalene

Response Ratio



$$\text{Response} = 1.789\text{e}+000 * \text{Amt} - 5.202\text{e}-002$$

Coef of Det (r^2) = 0.999397 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M

Calibration Table Last Updated: Thu Jan 27 06:04:38 2022