

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012623\
 Data File : VD075217.D
 Acq On : 26 Jan 2023 18:13
 Operator : KP/SY
 Sample : 01271-01RE
 Misc : 5.10G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-3-012423RE

Quant Time: Jan 27 00:26:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 19 01:32:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	3813	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	6765	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	6386	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	2756	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	2666	68.519	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 137.040%	
35) Dibromofluoromethane	7.810	113	2204	53.916	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.840%	
50) Toluene-d8	10.269	98	7899	49.076	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 98.160%	
62) 4-Bromofluorobenzene	12.581	95	2406	47.100	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 94.200%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012623\
Data File : VD075217.D
Acq On : 26 Jan 2023 18:13
Operator : KP/SY
Sample : 01271-01RE
Misc : 5.10G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-3-012423RE

Quant Time: Jan 27 00:26:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 19 01:32:27 2023
Response via : Initial Calibration

