

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012623\
 Data File : VD075218.D
 Acq On : 26 Jan 2023 18:40
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 27 00:26:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 19 01:32:27 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	72	0.00
2 T	Dichlorodifluoromethane	0.435	0.397	8.7	78	0.00
3 P	Chloromethane	0.646	0.625	3.3	74	0.00
4 C	Vinyl Chloride	0.693	0.677	2.3#	74	0.00
5 T	Bromomethane	0.471	0.434	7.9	69	0.00
6 T	Chloroethane	0.481	0.462	4.0	70	0.00
7 T	Trichlorofluoromethane	0.881	0.930	-5.6	78	0.00
8 T	Diethyl Ether	0.284	0.285	-0.4	68	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.544	0.586	-7.7	81	0.00
10 T	Methyl Iodide	0.732	0.698	4.6	62	0.00
11 T	Tert butyl alcohol	0.036	0.027	25.0	58	0.00
12 CM	1,1-Dichloroethene	0.573	0.568	0.9#	72	0.00
13 T	Acrolein	0.035	0.035	0.0	86	0.00
14 T	Allyl chloride	0.929	0.902	2.9	67	0.00
15 T	Acrylonitrile	0.120	0.127	-5.8	70	0.00
16 T	Acetone	0.135	0.132	2.2	59	0.00
17 T	Carbon Disulfide	1.857	1.821	1.9	71	0.00
18 T	Methyl Acetate	0.366	0.375	-2.5	69	0.00
19 T	Methyl tert-butyl Ether	1.196	1.232	-3.0	68	0.00
20 T	Methylene Chloride	0.849	0.749	11.8	68	0.00
21 T	trans-1,2-Dichloroethene	0.637	0.657	-3.1	72	0.00
22 T	Diisopropyl ether	1.729	1.849	-6.9	72	0.00
23 T	Vinyl Acetate	0.762	0.835	-9.6	76	0.00
24 P	1,1-Dichloroethane	1.087	1.180	-8.6	76	0.00
25 T	2-Butanone	0.163	0.168	-3.1	66	0.00
26 T	2,2-Dichloropropane	0.969	0.939	3.1	70	0.00
27 T	cis-1,2-Dichloroethene	0.695	0.733	-5.5	72	0.00
28 T	Bromochloromethane	0.297	0.293	1.3	71	0.00
29 T	Tetrahydrofuran	0.093	0.097	-4.3	69	0.00
30 C	Chloroform	1.084	1.151	-6.2#	73	0.00
31 T	Cyclohexane	1.023	1.026	-0.3	79	0.00
32 T	1,1,1-Trichloroethane	0.934	1.012	-8.4	77	0.00
33 S	1,2-Dichloroethane-d4	0.510	0.562	-10.2	75	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	71	0.00
35 S	Dibromofluoromethane	0.302	0.319	-5.6	71	0.00
36 T	1,1-Dichloropropene	0.485	0.526	-8.5	79	0.00
37 T	Ethyl Acetate	0.196	0.193	1.5	68	0.00
38 T	Carbon Tetrachloride	0.409	0.460	-12.5	81	0.00
39 T	Methylcyclohexane	0.598	0.631	-5.5	79	0.00
40 TM	Benzene	1.384	1.482	-7.1	73	0.00
41 T	Methacrylonitrile	0.118	0.111	5.9	69	0.00
42 TM	1,2-Dichloroethane	0.364	0.399	-9.6	74	0.00
43 T	Isopropyl Acetate	0.381	0.389	-2.1	69	0.00
44 TM	Trichloroethene	0.389	0.391	-0.5	69	0.00
45 C	1,2-Dichloropropane	0.341	0.365	-7.0#	73	0.00
46 T	Dibromomethane	0.180	0.191	-6.1	71	0.00
47 T	Bromodichloromethane	0.446	0.492	-10.3	74	0.00
48 T	Methyl methacrylate	0.186	0.197	-5.9	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012623\
 Data File : VD075218.D
 Acq On : 26 Jan 2023 18:40
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleID :
 VSTDCCC050

Quant Time: Jan 27 00:26:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 19 01:32:27 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	65	0.00
50 S	Toluene-d8	1.190	1.284	-7.9	74	0.00
51 T	4-Methyl-2-Pentanone	0.183	0.198	-8.2	74	0.00
52 CM	Toluene	0.875	0.927	-5.9#	72	0.00
53 T	t-1,3-Dichloropropene	0.422	0.444	-5.2	71	0.00
54 T	cis-1,3-Dichloropropene	0.516	0.557	-7.9	71	0.00
55 T	1,1,2-Trichloroethane	0.238	0.256	-7.6	72	0.00
56 T	Ethyl methacrylate	0.286	0.309	-8.0	70	0.00
57 T	1,3-Dichloropropane	0.409	0.455	-11.2	73	0.00
58 T	2-Chloroethyl Vinyl ether	0.098	0.109	-11.2	80	0.00
59 T	2-Hexanone	0.132	0.140	-6.1	69	0.00
60 T	Dibromochloromethane	0.287	0.310	-8.0	71	0.00
61 T	1,2-Dibromoethane	0.227	0.245	-7.9	72	0.00
62 S	4-Bromofluorobenzene	0.378	0.392	-3.7	70	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	71	0.00
64 T	Tetrachloroethene	0.357	0.367	-2.8	74	0.00
65 PM	Chlorobenzene	1.025	1.065	-3.9	72	0.00
66 T	1,1,1,2-Tetrachloroethane	0.345	0.371	-7.5	75	0.00
67 C	Ethyl Benzene	1.844	1.960	-6.3#	75	0.00
68 T	m/p-Xylenes	0.712	0.774	-8.7	77	0.00
69 T	o-Xylene	0.668	0.719	-7.6	73	0.00
70 T	Styrene	1.111	1.199	-7.9	74	0.00
71 P	Bromoform	0.178	0.202	-13.5	76	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
73 T	Isopropylbenzene	3.949	4.032	-2.1	78	0.00
74 T	N-amyl acetate	0.865	0.830	4.0	71	0.00
75 P	1,1,2,2-Tetrachloroethane	0.629	0.642	-2.1	74	0.00
76 T	1,2,3-Trichloropropane	0.423	0.426	-0.7	80	0.00
77 T	Bromobenzene	0.902	0.864	4.2	69	0.00
78 T	n-propylbenzene	4.853	4.956	-2.1	78	0.00
79 T	2-Chlorotoluene	2.669	2.674	-0.2	75	0.00
80 T	1,3,5-Trimethylbenzene	3.236	3.333	-3.0	76	0.00
81 T	trans-1,4-Dichloro-2-butene	0.183	0.175	4.4	71	0.00
82 T	4-Chlorotoluene	2.744	2.775	-1.1	76	0.00
83 T	tert-Butylbenzene	2.800	2.948	-5.3	80	0.00
84 T	1,2,4-Trimethylbenzene	3.178	3.248	-2.2	76	0.00
85 T	sec-Butylbenzene	4.253	4.385	-3.1	79	0.00
86 T	p-Isopropyltoluene	3.482	3.619	-3.9	79	0.00
87 T	1,3-Dichlorobenzene	1.749	1.748	0.1	74	0.00
88 T	1,4-Dichlorobenzene	1.720	1.733	-0.8	75	0.00
89 T	n-Butylbenzene	3.393	3.491	-2.9	79	0.00
90 T	Hexachloroethane	0.615	0.622	-1.1	78	0.00
91 T	1,2-Dichlorobenzene	1.496	1.515	-1.3	73	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.096	0.101	-5.2	73	0.00
93 T	1,2,4-Trichlorobenzene	0.982	0.919	6.4	68	0.00
94 T	Hexachlorobutadiene	0.481	0.472	1.9	74	0.00
95 T	Naphthalene	1.956	1.757	10.2	68	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012623\
Data File : VD075218.D
Acq On : 26 Jan 2023 18:40
Operator : KP/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jan 27 00:26:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 19 01:32:27 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.838	0.783	6.6	68	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6