

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD012623\  
 Data File : VD075212.D  
 Acq On : 26 Jan 2023 15:39  
 Operator : KP/SY  
 Sample : MDL06  
 Misc : 5.00G/5.00ml/MSVOA\_D/SOIL  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 MDL06

Manual Integrations  
 APPROVED

Reviewed By :Krupa Patel 01/27/2023  
 Supervised By :Mahesh Dadoda 01/27/2023

Quant Time: Jan 27 02:32:23 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_D\Method\82D011823S.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jan 27 02:32:07 2023  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 7.881  | 168  | 62041    | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 8.781  | 114  | 113722   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 11.587 | 117  | 100800   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 13.522 | 152  | 44236    | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 8.228  | 65    | 35633    | 56.285   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 50 - 163 | Recovery | =    | 112.580% |
| 35) Dibromofluoromethane    | 7.810  | 113   | 36342    | 52.885   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 54 - 147 | Recovery | =    | 105.780% |
| 50) Toluene-d8              | 10.275 | 98    | 139737   | 51.645   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 49 - 140 | Recovery | =    | 103.300% |
| 62) 4-Bromofluorobenzene    | 12.575 | 95    | 41653    | 48.506   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 25 - 144 | Recovery | =    | 97.020%  |

|                              |       |     |       |        |        |    |
|------------------------------|-------|-----|-------|--------|--------|----|
| Target Compounds             |       |     |       | Qvalue |        |    |
| 2) Dichlorodifluoromethane   | 1.934 | 85  | 1663  | 3.078  | ug/l   | 88 |
| 3) Chloromethane             | 2.146 | 50  | 2792  | 3.484  | ug/l   | 90 |
| 4) Vinyl Chloride            | 2.281 | 62  | 2392  | 2.782  | ug/l   | 96 |
| 5) Bromomethane              | 2.699 | 94  | 2168  | 3.707  | ug/l   | 95 |
| 6) Chloroethane              | 2.846 | 64  | 1847  | 3.098  | ug/l   | 99 |
| 7) Trichlorofluoromethane    | 3.181 | 101 | 3596  | 3.290  | ug/l   | 98 |
| 8) Diethyl Ether             | 3.599 | 74  | 938   | 2.663  | ug/l   | 91 |
| 9) 1,1,2-Trichlorotrifluo... | 3.963 | 101 | 2044  | 3.029  | ug/l # | 93 |
| 10) Methyl Iodide            | 4.175 | 142 | 1888  | 2.078  | ug/l # | 85 |
| 11) Tert butyl alcohol       | 5.052 | 59  | 677m  | 15.329 | ug/l   |    |
| 12) 1,1-Dichloroethene       | 3.946 | 96  | 2254  | 3.173  | ug/l   | 85 |
| 13) Acrolein                 | 3.799 | 56  | 636   | 14.826 | ug/l # | 64 |
| 14) Allyl chloride           | 4.569 | 41  | 3150  | 2.732  | ug/l   | 93 |
| 15) Acrylonitrile            | 5.269 | 53  | 1985  | 13.329 | ug/l   | 97 |
| 16) Acetone                  | 4.010 | 43  | 2436  | 14.589 | ug/l # | 79 |
| 17) Carbon Disulfide         | 4.269 | 76  | 6438  | 2.794  | ug/l # | 94 |
| 18) Methyl Acetate           | 4.558 | 43  | 1620  | 3.569  | ug/l # | 61 |
| 19) Methyl tert-butyl Ether  | 5.334 | 73  | 4228  | 2.849  | ug/l # | 52 |
| 20) Methylene Chloride       | 4.805 | 84  | 5920  | 5.623  | ug/l   | 95 |
| 21) trans-1,2-Dichloroethene | 5.310 | 96  | 2428  | 3.070  | ug/l # | 89 |
| 22) Diisopropyl ether        | 6.216 | 45  | 6112  | 2.848  | ug/l # | 93 |
| 23) Vinyl Acetate            | 6.163 | 43  | 10265 | 10.854 | ug/l # | 86 |
| 24) 1,1-Dichloroethane       | 6.116 | 63  | 4152  | 3.079  | ug/l # | 80 |
| 25) 2-Butanone               | 7.087 | 43  | 3213  | 15.921 | ug/l # | 77 |
| 26) 2,2-Dichloropropane      | 7.069 | 77  | 3830  | 3.186  | ug/l   | 83 |
| 27) cis-1,2-Dichloroethene   | 7.075 | 96  | 2668  | 3.095  | ug/l   | 94 |
| 28) Bromochloromethane       | 7.434 | 49  | 1072  | 2.912  | ug/l # | 81 |
| 29) Tetrahydrofuran          | 7.451 | 42  | 1650  | 14.335 | ug/l # | 80 |
| 30) Chloroform               | 7.593 | 83  | 4347  | 3.233  | ug/l   | 91 |
| 31) Cyclohexane              | 7.881 | 56  | 5451  | 4.293  | ug/l # | 88 |
| 32) 1,1,1-Trichloroethane    | 7.799 | 97  | 3517  | 3.036  | ug/l   | 98 |
| 36) 1,1-Dichloropropene      | 8.010 | 75  | 3220  | 2.916  | ug/l   | 92 |
| 37) Ethyl Acetate            | 7.181 | 43  | 1157m | 2.601  | ug/l   |    |
| 38) Carbon Tetrachloride     | 7.998 | 117 | 2441  | 2.626  | ug/l   | 87 |
| 39) Methylcyclohexane        | 9.281 | 83  | 3942  | 2.897  | ug/l # | 86 |
| 40) Benzene                  | 8.251 | 78  | 8903  | 2.828  | ug/l   | 99 |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD012623\  
 Data File : VD075212.D  
 Acq On : 26 Jan 2023 15:39  
 Operator : KP/SY  
 Sample : MDL06  
 Misc : 5.00G/5.00ml/MSVOA\_D/SOIL  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 MDL06

Manual Integrations  
 APPROVED

Reviewed By :Krupa Patel 01/27/2023  
 Supervised By :Mahesh Dadoda 01/27/2023

Quant Time: Jan 27 02:32:23 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_D\Method\82D011823S.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jan 27 02:32:07 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|--------|--------|----------|
| 41) Methacrylonitrile         | 7.393  | 41   | 385m     | 1.438  | ug/l   |          |
| 42) 1,2-Dichloroethane        | 8.328  | 62   | 2426     | 2.933  | ug/l # | 74       |
| 43) Isopropyl Acetate         | 8.363  | 43   | 2337     | 2.700  | ug/l # | 83       |
| 44) Trichloroethene           | 9.034  | 130  | 2634     | 2.975  | ug/l   | 97       |
| 45) 1,2-Dichloropropane       | 9.310  | 63   | 2114     | 2.726  | ug/l   | 93       |
| 46) Dibromomethane            | 9.393  | 93   | 1164     | 2.846  | ug/l # | 82       |
| 47) Bromodichloromethane      | 9.581  | 83   | 2832     | 2.791  | ug/l # | 77       |
| 48) Methyl methacrylate       | 9.387  | 41   | 1135     | 2.681  | ug/l   | 93       |
| 51) 4-Methyl-2-Pentanone      | 10.157 | 43   | 5689     | 13.686 | ug/l   | 95       |
| 52) Toluene                   | 10.340 | 92   | 5800     | 2.915  | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 10.557 | 75   | 2583     | 2.690  | ug/l   | 95       |
| 54) cis-1,3-Dichloropropene   | 10.022 | 75   | 3099     | 2.642  | ug/l   | 92       |
| 55) 1,1,2-Trichloroethane     | 10.734 | 97   | 1313     | 2.426  | ug/l   | 95       |
| 56) Ethyl methacrylate        | 10.604 | 69   | 1555     | 2.389  | ug/l # | 88       |
| 57) 1,3-Dichloropropane       | 10.881 | 76   | 2657     | 2.856  | ug/l   | 91       |
| 58) 2-Chloroethyl Vinyl ether | 9.869  | 63   | 3134     | 14.069 | ug/l   | 98       |
| 59) 2-Hexanone                | 10.928 | 43   | 3763     | 12.520 | ug/l   | 95       |
| 60) Dibromochloromethane      | 11.075 | 129  | 1673     | 2.564  | ug/l   | 83       |
| 61) 1,2-Dibromoethane         | 11.175 | 107  | 1460     | 2.823  | ug/l   | 93       |
| 64) Tetrachloroethene         | 10.810 | 164  | 2488     | 3.456  | ug/l # | 85       |
| 65) Chlorobenzene             | 11.610 | 112  | 6124     | 2.965  | ug/l # | 88       |
| 66) 1,1,1,2-Tetrachloroethane | 11.681 | 131  | 2011     | 2.895  | ug/l   | 96       |
| 67) Ethyl Benzene             | 11.687 | 91   | 10428    | 2.805  | ug/l   | 93       |
| 68) m/p-Xylenes               | 11.798 | 106  | 7524     | 5.239  | ug/l   | 95       |
| 69) o-Xylene                  | 12.128 | 106  | 3793     | 2.815  | ug/l   | 91       |
| 70) Styrene                   | 12.139 | 104  | 6094     | 2.721  | ug/l   | 98       |
| 71) Bromoform                 | 12.298 | 173  | 854      | 2.381  | ug/l # | 82       |
| 73) Isopropylbenzene          | 12.422 | 105  | 9727     | 2.784  | ug/l   | 97       |
| 74) N-amyl acetate            | 12.239 | 43   | 1829     | 2.390  | ug/l   | 97       |
| 75) 1,1,2,2-Tetrachloroethane | 12.675 | 83   | 1700     | 3.054  | ug/l # | 88       |
| 76) 1,2,3-Trichloropropane    | 12.728 | 75   | 1270m    | 3.390  | ug/l   |          |
| 77) Bromobenzene              | 12.710 | 156  | 2273     | 2.848  | ug/l   | 99       |
| 78) n-propylbenzene           | 12.769 | 91   | 13191    | 3.072  | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 12.851 | 91   | 7084     | 3.000  | ug/l   | 100      |
| 80) 1,3,5-Trimethylbenzene    | 12.910 | 105  | 8188     | 2.860  | ug/l   | 98       |
| 81) trans-1,4-Dichloro-2-b... | 12.475 | 75   | 277m     | 1.715  | ug/l   |          |
| 82) 4-Chlorotoluene           | 12.951 | 91   | 6753     | 2.781  | ug/l   | 95       |
| 83) tert-Butylbenzene         | 13.169 | 119  | 7231     | 2.919  | ug/l   | 97       |
| 84) 1,2,4-Trimethylbenzene    | 13.216 | 105  | 7200     | 2.561  | ug/l   | 95       |
| 85) sec-Butylbenzene          | 13.351 | 105  | 10803    | 2.871  | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 13.463 | 119  | 7956     | 2.583  | ug/l   | 98       |
| 87) 1,3-Dichlorobenzene       | 13.463 | 146  | 4475     | 2.893  | ug/l   | 95       |
| 88) 1,4-Dichlorobenzene       | 13.539 | 146  | 4983     | 3.274  | ug/l   | 93       |
| 89) n-Butylbenzene            | 13.792 | 91   | 8685     | 2.893  | ug/l   | 98       |
| 90) Hexachloroethane          | 14.063 | 117  | 1706     | 3.136  | ug/l   | 86       |
| 91) 1,2-Dichlorobenzene       | 13.839 | 146  | 4085     | 3.086  | ug/l   | 99       |
| 92) 1,2-Dibromo-3-Chloropr... | 14.451 | 75   | 133      | 1.562  | ug/l # | 45       |
| 93) 1,2,4-Trichlorobenzene    | 15.104 | 180  | 2367     | 2.723  | ug/l   | 95       |
| 94) Hexachlorobutadiene       | 15.210 | 225  | 1115     | 2.618  | ug/l   | 95       |
| 95) Naphthalene               | 15.339 | 128  | 4930     | 2.849  | ug/l   | 97       |
| 96) 1,2,3-Trichlorobenzene    | 15.528 | 180  | 2078     | 2.802  | ug/l   | 97       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD012623\  
Data File : VD075212.D  
Acq On : 26 Jan 2023 15:39  
Operator : KP/SY  
Sample : MDL06  
Misc : 5.00G/5.00ml/MSVOA\_D/SOIL  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06


Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023

Quant Time: Jan 27 02:32:23 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_D\Method\82D011823S.M  
Quant Title : SW846 8260  
QLast Update : Fri Jan 27 02:32:07 2023  
Response via : Initial Calibration

| Compound                                                             | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------------------------------------------------|------|------|----------|------|-------|----------|
| -----                                                                |      |      |          |      |       |          |
| (#)=qualifier out of range (m)=manual integration (+)=signals summed |      |      |          |      |       |          |





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.881 min Scan# 1069

Delta R.T. 0.006 min

Lab File: VD075212.D

Acq: 26 Jan 2023 15:39

Instrument :

MSVOA\_D

ClientSampleId :

MDL06

Tgt Ion:168 Resp: 6204

Ion Ratio Lower Upper

168 100





99 57.9 45.2 67.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 01/27/2023

Supervised By :Mahesh Dadoda 01/27/2023



#2

Dichlorodifluoromethane

Concen: 3.078 ug/l

RT: 1.934 min Scan# 59

Delta R.T. 0.000 min

Lab File: VD075212.D




Acq: 26 Jan 2023 15:39


Tgt Ion: 85 Resp: 1663

Ion Ratio Lower Upper

85 100


87 24.4 15.7 47.0





#3  
Chloromethane  
Concen: 3.484 ug/l  
RT: 2.146 min Scan# 91  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06




Tgt Ion: 50 Resp: 2792  
Ion Ratio Lower Upper  
50 100  
52 39.6 27.0 40.6

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023




#4  
Vinyl Chloride  
Concen: 2.782 ug/l  
RT: 2.281 min Scan# 118  
Delta R.T. -0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Tgt Ion: 62 Resp: 2392  
Ion Ratio Lower Upper  
62 100  
64 28.8 24.8 37.2





#5  
Bromomethane  
Concen: 3.707 ug/l  
RT: 2.699 min Scan# 11  
Delta R.T. 0.012 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06




Tgt Ion: 94 Resp: 2168  
Ion Ratio Lower Upper  
94 100  
96 88.7 74.8 112.2

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023




#6  
Chloroethane  
Concen: 3.098 ug/l  
RT: 2.846 min Scan# 214  
Delta R.T. 0.012 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39



Tgt Ion: 64 Resp: 1847  
Ion Ratio Lower Upper  
64 100  
66 31.1 25.4 38.2









#7  
Trichlorofluoromethane  
Concen: 3.290 ug/l  
RT: 3.181 min Scan# 21  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06

Tgt Ion:101 Resp: 3590  
Ion Ratio Lower Upper  
101 100  
103 66.9 52.2 78.4




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#8  
Diethyl Ether  
Concen: 2.663 ug/l  
RT: 3.599 min Scan# 342  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Tgt Ion: 74 Resp: 938  
Ion Ratio Lower Upper  
74 100  
45 90.4 49.8 149.3









#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 3.029 ug/l  
RT: 3.963 min Scan# 404  
Delta R.T. -0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06

Tgt Ion:101 Resp: 204  
Ion Ratio Lower Upper  
101 100  
85 57.4 36.9 55.3  
151 73.4 58.9 88.3




Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#10  
Methyl Iodide  
Concen: 2.078 ug/l  
RT: 4.175 min Scan# 440  
Delta R.T. 0.011 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39


Tgt Ion:142 Resp: 1888  
Ion Ratio Lower Upper  
142 100  
127 36.5 33.6 50.4  
141 0.0 11.1 16.7#





#11  
Tert butyl alcohol  
Concen: 15.329 ug/l m  
RT: 5.052 min Scan# 589  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06




Tgt Ion: 59 Resp: 67  
Ion Ratio Lower Upper  
59 100  
57 0.0 2.9 4.3

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023




#12  
1,1-Dichloroethene  
Concen: 3.173 ug/l  
RT: 3.946 min Scan# 401  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Tgt Ion: 96 Resp: 2254  
Ion Ratio Lower Upper  
96 100  
61 132.2 126.8 190.2  
98 61.7 50.4 75.6





#13  
Acrolein  
Concen: 14.826 ug/l  
RT: 3.799 min Scan# 31  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06




Tgt Ion: 56 Resp: 630  
Ion Ratio Lower Upper  
56 100  
55 37.4 53.0 79.6

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023




#14  
Allyl chloride  
Concen: 2.732 ug/l  
RT: 4.569 min Scan# 507  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Tgt Ion: 41 Resp: 3150  
Ion Ratio Lower Upper  
41 100  
39 76.9 55.4 83.0  
76 39.5 30.2 45.2





#15  
Acrylonitrile  
Concen: 13.329 ug/l  
RT: 5.269 min Scan# 61  
Delta R.T. 0.017 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06



Tgt Ion: 53 Resp: 198  
Ion Ratio Lower Upper  
53 100  
52 81.8 67.0 100.6  
51 34.4 30.1 45.1




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#16  
Acetone  
Concen: 14.589 ug/l  
RT: 4.010 min Scan# 412  
Delta R.T. -0.012 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Tgt Ion: 43 Resp: 2436  
Ion Ratio Lower Upper  
43 100  
58 20.7 26.0 39.0#









#17  
Carbon Disulfide  
Concen: 2.794 ug/l  
RT: 4.269 min Scan# 41  
Delta R.T. -0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06

Tgt Ion: 76 Resp: 6438  
Ion Ratio Lower Upper  
76 100  
78 7.0 7.4 11.0




Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#18  
Methyl Acetate  
Concen: 3.569 ug/l  
RT: 4.558 min Scan# 505  
Delta R.T. -0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Tgt Ion: 43 Resp: 1620  
Ion Ratio Lower Upper  
43 100  
74 4.3 18.5 27.7









#19  
Methyl tert-butyl Ether  
Concen: 2.849 ug/l  
RT: 5.334 min Scan# 61  
Delta R.T. 0.011 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06

Tgt Ion: 73 Resp: 4228  
Ion Ratio Lower Upper  
73 100  
57 0.0 18.5 27.7




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#20  
Methylene Chloride  
Concen: 5.623 ug/l  
RT: 4.805 min Scan# 547  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39


Tgt Ion: 84 Resp: 5920  
Ion Ratio Lower Upper  
84 100  
49 123.5 97.0 145.6  
51 34.1 30.9 46.3  
86 60.8 54.0 81.0





#21  
trans-1,2-Dichloroethene  
Concen: 3.070 ug/l  
RT: 5.310 min Scan# 610  
Delta R.T. -0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06



Tgt Ion: 96 Resp: 2428  
Ion Ratio Lower Upper  
96 100  
61 142.8 105.7 158.5  
98 50.3 50.4 75.6




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#22  
Diisopropyl ether  
Concen: 2.848 ug/l  
RT: 6.216 min Scan# 787  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39


Tgt Ion: 45 Resp: 6112  
Ion Ratio Lower Upper  
45 100  
43 50.1 45.3 67.9  
87 28.1 22.1 33.1  
59 15.2 9.1 13.7





#23  
 Vinyl Acetate  
 Concen: 10.854 ug/l  
 RT: 6.163 min Scan# 71  
 Delta R.T. 0.006 min  
 Lab File: VD075212.D  
 Acq: 26 Jan 2023 15:39




Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 MDL06




Tgt Ion: 43 Resp: 1026  
 Ion Ratio Lower Upper  
 43 100  
 86 17.0 9.3 13.9

Manual Integrations  
 APPROVED



Reviewed By :Krupa Patel 01/27/2023  
 Supervised By :Mahesh Dadoda 01/27/2023




#24  
 1,1-Dichloroethane  
 Concen: 3.079 ug/l  
 RT: 6.116 min Scan# 770  
 Delta R.T. -0.000 min  
 Lab File: VD075212.D  
 Acq: 26 Jan 2023 15:39



Tgt Ion: 63 Resp: 4152  
 Ion Ratio Lower Upper  
 63 100  
 98 0.0 4.0 12.2#  
 100 0.0 2.5 7.4#









#25  
2-Butanone  
Concen: 15.921 ug/l  
RT: 7.087 min Scan# 91  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06

Tgt Ion: 43 Resp: 321  
Ion Ratio Lower Upper  
43 100  
72 36.5 19.8 29.8




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#26  
2,2-Dichloropropane  
Concen: 3.186 ug/l  
RT: 7.069 min Scan# 932  
Delta R.T. -0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Tgt Ion: 77 Resp: 3830  
Ion Ratio Lower Upper  
77 100  
97 15.7 12.0 36.0









#27  
cis-1,2-Dichloroethene  
Concen: 3.095 ug/l  
RT: 7.075 min Scan# 91  
Delta R.T. -0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06

Tgt Ion: 96 Resp: 2668  
Ion Ratio Lower Upper  
96 100  
61 147.5 0.0 287.4  
98 55.8 0.0 129.8




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#28  
Bromochloromethane  
Concen: 2.912 ug/l  
RT: 7.434 min Scan# 994  
Delta R.T. 0.012 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Tgt Ion: 49 Resp: 1072  
Ion Ratio Lower Upper  
49 100  
129 0.0 0.0 2.6  
130 62.2 63.6 95.4#





#29

Tetrahydrofuran

Concen: 14.335 ug/l

RT: 7.451 min Scan# 996

Delta R.T. 0.006 min

Lab File: VD075212.D

Acq: 26 Jan 2023 15:39

Instrument :

MSVOA\_D

ClientSampleId :

MDL06

Tgt Ion: 42 Resp: 1650





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 42  | 100   |       |       |
| 72  | 27.9  | 33.8  | 50.8  |
| 71  | 29.1  | 31.6  | 47.4  |

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 01/27/2023

Supervised By :Mahesh Dadoda 01/27/2023



#30

Chloroform

Concen: 3.233 ug/l

RT: 7.593 min Scan# 1021




Delta R.T. -0.006 min


Lab File: VD075212.D

Acq: 26 Jan 2023 15:39

Tgt Ion: 83 Resp: 4347


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 83  | 100   |       |       |
| 85  | 67.4  | 48.5  | 72.7  |





#31  
Cyclohexane  
Concen: 4.293 ug/l  
RT: 7.881 min Scan# 1070  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06



Tgt Ion: 56 Resp: 545  
Ion Ratio Lower Upper  
56 100  
69 52.6 26.6 40.0  
84 89.1 74.1 111.1




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#32  
1,1,1-Trichloroethane  
Concen: 3.036 ug/l  
RT: 7.799 min Scan# 1056  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Tgt Ion: 97 Resp: 3517  
Ion Ratio Lower Upper  
97 100  
99 64.2 51.7 77.5  
61 42.2 36.3 54.5





#33

1,2-Dichloroethane-d4

Concen: 56.285 ug/l

RT: 8.228 min Scan# 1129

Delta R.T. -0.000 min

Lab File: VD075212.D

Acq: 26 Jan 2023 15:39

Instrument :

MSVOA\_D

ClientSampleId :

MDL06

Tgt Ion: 65 Resp: 3563

Ion Ratio Lower Upper

65 100





67 54.4 0.0 105.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 01/27/2023

Supervised By :Mahesh Dadoda 01/27/2023



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 1223

Delta R.T. -0.000 min

Lab File: VD075212.D

Acq: 26 Jan 2023 15:39




Tgt Ion:114 Resp: 113722

Ion Ratio Lower Upper

114 100

63 19.4 0.0 40.2

88 14.6 0.0 32.2





#35

Dibromofluoromethane

Concen: 52.885 ug/l

RT: 7.810 min Scan# 1058

Delta R.T. -0.000 min

Lab File: VD075212.D

Acq: 26 Jan 2023 15:39

Instrument :

MSVOA\_D

ClientSampleId :

MDL06

Tgt Ion: 113 Resp: 3634





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 113 | 100   |       |       |
| 111 | 102.5 | 80.1  | 120.1 |
| 192 | 18.0  | 15.4  | 23.0  |

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 01/27/2023

Supervised By :Mahesh Dadoda 01/27/2023



#36

1,1-Dichloropropene

Concen: 2.916 ug/l

RT: 8.010 min Scan# 1092




Delta R.T. -0.000 min

Lab File: VD075212.D

Acq: 26 Jan 2023 15:39

Tgt Ion: 75 Resp: 3220

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 75  | 100   |       |       |
| 110 | 32.4  | 18.1  | 54.4  |
| 77  | 36.4  | 25.1  | 37.7  |





#37

Ethyl Acetate

Concen: 2.601 ug/l m

RT: 7.181 min Scan# 91

Delta R.T. 0.012 min

Lab File: VD075212.D


Acq: 26 Jan 2023 15:39

Instrument :

MSVOA\_D

ClientSampleId :

MDL06






|           |       |       |      |
|-----------|-------|-------|------|
| Tgt Ion:  | 43    | Resp: | 115  |
| Ion Ratio | Lower | Upper |      |
| 43        | 100   |       |      |
| 61        | 0.0   | 11.7  | 17.5 |
| 70        | 0.0   | 8.8   | 13.2 |

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 01/27/2023

Supervised By :Mahesh Dadoda 01/27/2023



#38

Carbon Tetrachloride

Concen: 2.626 ug/l




RT: 7.998 min Scan# 1090


Delta R.T. 0.006 min

Lab File: VD075212.D

Acq: 26 Jan 2023 15:39

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 117   | Resp: | 2441  |
| Ion Ratio | Lower | Upper |       |
| 117       | 100   |       |       |
| 119       | 76.4  | 73.6  | 110.4 |
| 121       | 29.6  | 25.4  | 38.0  |





#39

Methylcyclohexane

Concen: 2.897 ug/l

RT: 9.281 min Scan# 11308

Delta R.T. 0.006 min

Lab File: VD075212.D

Acq: 26 Jan 2023 15:39

Instrument :

MSVOA\_D

ClientSampleId :

MDL06





|             |             |
|-------------|-------------|
| Tgt Ion: 83 | Resp: 394   |
| Ion Ratio   | Lower Upper |
| 83 100      |             |
| 55 59.8     | 60.0 90.0   |
| 98 43.5     | 39.0 58.4   |

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 01/27/2023

Supervised By :Mahesh Dadoda 01/27/2023



#40

Benzene

Concen: 2.828 ug/l




RT: 8.251 min Scan# 1133


Delta R.T. -0.000 min

Lab File: VD075212.D

Acq: 26 Jan 2023 15:39


|             |             |
|-------------|-------------|
| Tgt Ion: 78 | Resp: 8903  |
| Ion Ratio   | Lower Upper |
| 78 100      |             |
| 77 25.3     | 19.8 29.6   |





#41  
Methacrylonitrile  
Concen: 1.438 ug/l m  
RT: 7.393 min Scan# 987  
Delta R.T. -0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06



Tgt Ion: 41 Resp: 385  
Ion Ratio Lower Upper  
41 100  
39 0.0 57.8 86.6  
67 74.0 60.5 90.7  
52 0.0 25.4 38.0#




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#42  
1,2-Dichloroethane  
Concen: 2.933 ug/l  
RT: 8.328 min Scan# 1146  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39


Tgt Ion: 62 Resp: 2426  
Ion Ratio Lower Upper  
62 100  
98 0.0 0.0 19.2





#43  
Isopropyl Acetate  
Concen: 2.700 ug/l  
RT: 8.363 min Scan# 1152  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06




Tgt Ion: 43 Resp: 233  
Ion Ratio Lower Upper  
43 100  
61 23.0 25.0 37.6  
87 5.2 10.6 16.0

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023




#44  
Trichloroethene  
Concen: 2.975 ug/l  
RT: 9.034 min Scan# 1266  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39



Tgt Ion:130 Resp: 2634  
Ion Ratio Lower Upper  
130 100  
95 91.1 0.0 188.6





#45

1,2-Dichloropropane

Concen: 2.726 ug/l

RT: 9.310 min Scan# 1114

Delta R.T. 0.006 min

Lab File: VD075212.D

Acq: 26 Jan 2023 15:39

Instrument :

MSVOA\_D

ClientSampleId :

MDL06

Tgt Ion: 63 Resp: 2114

Ion Ratio Lower Upper

63 100





65 27.3 25.1 37.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 01/27/2023

Supervised By :Mahesh Dadoda 01/27/2023



#46

Dibromomethane

Concen: 2.846 ug/l

RT: 9.393 min Scan# 1327

Delta R.T. -0.006 min

Lab File: VD075212.D

Acq: 26 Jan 2023 15:39




Tgt Ion: 93 Resp: 1164


Ion Ratio Lower Upper

93 100

95 74.5 68.3 102.5

174 73.5 77.4 116.0#





#47

Bromodichloromethane

Concen: 2.791 ug/l

RT: 9.581 min Scan# 1360

Delta R.T. -0.006 min

Lab File: VD075212.D

Acq: 26 Jan 2023 15:39

Instrument :

MSVOA\_D

ClientSampleId :

MDL06

Tgt Ion: 83 Resp: 283





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 83  | 100   |       |       |
| 85  | 47.8  | 53.0  | 79.4  |
| 127 | 0.0   | 7.0   | 10.4  |

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 01/27/2023

Supervised By :Mahesh Dadoda 01/27/2023



#48

Methyl methacrylate

Concen: 2.681 ug/l

RT: 9.387 min Scan# 1326




Delta R.T. 0.006 min


Lab File: VD075212.D

Acq: 26 Jan 2023 15:39

Tgt Ion: 41 Resp: 1135


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 41  | 100   |       |       |
| 69  | 76.2  | 67.8  | 101.6 |
| 39  | 54.7  | 46.0  | 69.0  |





#50  
Toluene-d8  
Concen: 51.645 ug/l  
RT: 10.275 min Scan# 1476  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06




Tgt Ion: 98 Resp: 13973  
Ion Ratio Lower Upper  
98 100  
100 65.1 52.7 79.1

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023




#51  
4-Methyl-2-Pentanone  
Concen: 13.686 ug/l  
RT: 10.157 min Scan# 1457  
Delta R.T. -0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39



Tgt Ion: 43 Resp: 5689  
Ion Ratio Lower Upper  
43 100  
58 36.2 31.4 47.0









#52  
Toluene  
Concen: 2.915 ug/l  
RT: 10.340 min Scan# 1488  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06

Tgt Ion: 92 Resp: 5800  
Ion Ratio Lower Upper  
92 100  
91 172.8 137.0 205.4




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#53  
t-1,3-Dichloropropene  
Concen: 2.690 ug/l  
RT: 10.557 min Scan# 1525  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39


Tgt Ion: 75 Resp: 2583  
Ion Ratio Lower Upper  
75 100  
77 28.6 25.0 37.6





#54  
cis-1,3-Dichloropropene  
Concen: 2.642 ug/l  
RT: 10.022 min Scan# 1433  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06



Tgt Ion: 75 Resp: 3099  
Ion Ratio Lower Upper  
75 100  
77 32.0 25.1 37.7  
39 55.8 37.9 56.9




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#55  
1,1,2-Trichloroethane  
Concen: 2.426 ug/l  
RT: 10.734 min Scan# 1555  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39


Tgt Ion: 97 Resp: 1313  
Ion Ratio Lower Upper  
97 100  
83 83.7 69.4 104.0  
85 49.5 46.0 69.0  
99 65.6 51.3 76.9





#56  
Ethyl methacrylate  
Concen: 2.389 ug/l  
RT: 10.604 min Scan# 1533  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06



Tgt Ion: 69 Resp: 155  
Ion Ratio Lower Upper  
69 100  
41 77.6 55.7 83.5  
39 48.1 31.4 47.0




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#57  
1,3-Dichloropropane  
Concen: 2.856 ug/l  
RT: 10.881 min Scan# 1580  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Tgt Ion: 76 Resp: 2657  
Ion Ratio Lower Upper  
76 100  
78 27.7 26.2 39.2





#58

2-Chloroethyl Vinyl ether

Concen: 14.069 ug/l

RT: 9.869 min Scan# 1408

Delta R.T. -0.000 min

Lab File: VD075212.D

Acq: 26 Jan 2023 15:39

Instrument :

MSVOA\_D

ClientSampleId :

MDL06

Tgt Ion: 63 Resp: 3134

Ion Ratio Lower Upper

63 100





106 27.2 22.6 33.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 01/27/2023

Supervised By :Mahesh Dadoda 01/27/2023



#59

2-Hexanone

Concen: 12.520 ug/l

RT: 10.928 min Scan# 1588

Delta R.T. 0.006 min

Lab File: VD075212.D




Acq: 26 Jan 2023 15:39


Tgt Ion: 43 Resp: 3763

Ion Ratio Lower Upper

43 100

58 50.3 26.8 80.4









#60  
Dibromochloromethane  
Concen: 2.564 ug/l  
RT: 11.075 min Scan# 1613  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06

Tgt Ion:129 Resp: 1673  
Ion Ratio Lower Upper  
129 100  
127 92.4 38.9 116.6




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#61  
1,2-Dibromoethane  
Concen: 2.823 ug/l  
RT: 11.175 min Scan# 1630  
Delta R.T. -0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39


Tgt Ion:107 Resp: 1460  
Ion Ratio Lower Upper  
107 100  
109 87.1 75.2 112.8





#62  
4-Bromofluorobenzene  
Concen: 48.506 ug/l  
RT: 12.575 min Scan# 1868  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06




Tgt Ion: 95 Resp: 4165  
Ion Ratio Lower Upper  
95 100  
174 80.8 0.0 161.2  
176 78.6 0.0 160.2

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023




#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.587 min Scan# 1700  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Tgt Ion:117 Resp: 100800  
Ion Ratio Lower Upper  
117 100  
82 52.3 39.6 59.4  
119 32.1 24.6 37.0





#64  
Tetrachloroethene  
Concen: 3.456 ug/l  
RT: 10.810 min Scan# 1568  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06



Tgt Ion:164 Resp: 2488  
Ion Ratio Lower Upper  
164 100  
166 114.3 105.3 157.9  
129 88.0 76.2 114.4  
131 68.7 71.3 106.9#




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#65  
Chlorobenzene  
Concen: 2.965 ug/l  
RT: 11.610 min Scan# 1704  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Tgt Ion:112 Resp: 6124  
Ion Ratio Lower Upper  
112 100  
114 37.3 24.7 37.1#









#66  
1,1,1,2-Tetrachloroethane  
Concen: 2.895 ug/l  
RT: 11.681 min Scan# 1716  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06

Tgt Ion:131 Resp: 201  
Ion Ratio Lower Upper  
131 100  
133 97.5 49.0 147.2  
119 77.1 34.8 104.5




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#67  
Ethyl Benzene  
Concen: 2.805 ug/l  
RT: 11.687 min Scan# 1717  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Tgt Ion: 91 Resp: 10428  
Ion Ratio Lower Upper  
91 100  
106 35.4 25.0 37.6









#68  
m/p-Xylenes  
Concen: 5.239 ug/l  
RT: 11.798 min Scan# 1736  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06

Tgt Ion:106 Resp: 7524  
Ion Ratio Lower Upper  
106 100  
91 208.3 160.3 240.5




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#69  
o-Xylene  
Concen: 2.815 ug/l  
RT: 12.128 min Scan# 1792  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Tgt Ion:106 Resp: 3793  
Ion Ratio Lower Upper  
106 100  
91 192.6 103.6 310.9









#70  
Styrene  
Concen: 2.721 ug/l  
RT: 12.139 min Scan# 1794  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06

Tgt Ion:104 Resp: 6094  
Ion Ratio Lower Upper  
104 100  
78 47.6 38.9 58.3  
103 52.1 43.6 65.4




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#71  
Bromoform  
Concen: 2.381 ug/l  
RT: 12.298 min Scan# 1821  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39


Tgt Ion:173 Resp: 854  
Ion Ratio Lower Upper  
173 100  
175 38.5 25.4 76.0  
254 0.0 0.0 0.0





#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.522 min Scan# 2029  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06




Tgt Ion:152 Resp: 44230  
Ion Ratio Lower Upper  
152 100  
115 59.4 28.8 86.4  
150 158.5 0.0 348.6

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023




#73  
Isopropylbenzene  
Concen: 2.784 ug/l  
RT: 12.422 min Scan# 1842  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39



Tgt Ion:105 Resp: 9727  
Ion Ratio Lower Upper  
105 100  
120 25.8 13.7 41.1





#74  
N-amy1 acetate  
Concen: 2.390 ug/l  
RT: 12.239 min Scan# 1829  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39





Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06

Tgt Ion: 43 Resp: 1829

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 43  | 100   |       |       |
| 70  | 43.1  | 32.6  | 49.0  |
| 55  | 21.5  | 17.8  | 26.8  |
| 61  | 21.7  | 18.7  | 28.1  |

Manual Integrations  
APPROVED




Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023




#75  
1,1,2,2-Tetrachloroethane  
Concen: 3.054 ug/l  
RT: 12.675 min Scan# 1885  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Tgt Ion: 83 Resp: 1700

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 83  | 100   |       |       |
| 131 | 0.0   | 4.9   | 14.6# |
| 85  | 71.4  | 31.9  | 95.7  |









#76  
1,2,3-Trichloropropane  
Concen: 3.390 ug/l m  
RT: 12.728 min Scan# 1894  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06

Tgt Ion: 75 Resp: 1270  
Ion Ratio Lower Upper  
75 100  
77 0.0 0.0 0.0




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#77  
Bromobenzene  
Concen: 2.848 ug/l  
RT: 12.710 min Scan# 1891  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Tgt Ion:156 Resp: 2273  
Ion Ratio Lower Upper  
156 100  
77 172.3 85.6 256.8  
158 95.5 47.3 141.9









#78  
n-propylbenzene  
Concen: 3.072 ug/l  
RT: 12.769 min Scan# 1900  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06

Tgt Ion: 91 Resp: 1319  
Ion Ratio Lower Upper  
91 100  
120 22.8 11.7 35.0




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#79  
2-Chlorotoluene  
Concen: 3.000 ug/l  
RT: 12.851 min Scan# 1915  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Tgt Ion: 91 Resp: 7084  
Ion Ratio Lower Upper  
91 100  
126 34.8 17.4 52.3





#80

1,3,5-Trimethylbenzene

Concen: 2.860 ug/l

RT: 12.910 min Scan# 1924

Delta R.T. 0.006 min

Lab File: VD075212.D

Acq: 26 Jan 2023 15:39

Instrument :

MSVOA\_D

ClientSampleId :

MDL06

Tgt Ion:105 Resp: 818

Ion Ratio Lower Upper

105 100





120 50.8 24.6 73.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 01/27/2023

Supervised By :Mahesh Dadoda 01/27/2023



#81

trans-1,4-Dichloro-2-butene

Concen: 1.715 ug/l m

RT: 12.475 min Scan# 1851

Delta R.T. 0.006 min

Lab File: VD075212.D

Acq: 26 Jan 2023 15:39




Tgt Ion: 75 Resp: 277


Ion Ratio Lower Upper

75 100

53 0.0 70.5 105.7#

89 0.0 34.6 52.0#









#82  
4-Chlorotoluene  
Concen: 2.781 ug/l  
RT: 12.951 min Scan# 1932  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06

Tgt Ion: 91 Resp: 675  
Ion Ratio Lower Upper  
91 100  
126 37.5 17.3 52.0




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#83  
tert-Butylbenzene  
Concen: 2.919 ug/l  
RT: 13.169 min Scan# 1969  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39


Tgt Ion:119 Resp: 7231  
Ion Ratio Lower Upper  
119 100  
91 64.4 32.9 98.6  
134 23.2 13.3 39.8





#84  
1,2,4-Trimethylbenzene  
Concen: 2.561 ug/l  
RT: 13.216 min Scan# 1977  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06




Tgt Ion:105 Resp: 7200  
Ion Ratio Lower Upper  
105 100  
120 49.7 23.2 69.5

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023




#85  
sec-Butylbenzene  
Concen: 2.871 ug/l  
RT: 13.351 min Scan# 2000  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Tgt Ion:105 Resp: 10803  
Ion Ratio Lower Upper  
105 100  
134 20.1 10.4 31.2





#86  
p-Isopropyltoluene  
Concen: 2.583 ug/l  
RT: 13.463 min Scan# 2019  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06




Tgt Ion:119 Resp: 7950  
Ion Ratio Lower Upper  
119 100  
134 25.8 13.3 39.8  
91 24.8 11.9 35.5

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023




#87  
1,3-Dichlorobenzene  
Concen: 2.893 ug/l  
RT: 13.463 min Scan# 2019  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39



Tgt Ion:146 Resp: 4475  
Ion Ratio Lower Upper  
146 100  
111 38.4 18.9 56.9  
148 67.9 31.1 93.3









#88  
1,4-Dichlorobenzene  
Concen: 3.274 ug/l  
RT: 13.539 min Scan# 2032  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06

Tgt Ion:146 Resp: 498  
Ion Ratio Lower Upper  
146 100  
111 49.7 19.7 59.1  
148 63.9 31.9 95.5




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#89  
n-Butylbenzene  
Concen: 2.893 ug/l  
RT: 13.792 min Scan# 2075  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Tgt Ion: 91 Resp: 8685  
Ion Ratio Lower Upper  
91 100  
92 51.6 26.8 80.3  
134 26.2 13.3 39.8









#90  
Hexachloroethane  
Concen: 3.136 ug/l  
RT: 14.063 min Scan# 2120  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39

Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06

Tgt Ion:117 Resp: 1700  
Ion Ratio Lower Upper  
117 100  
201 58.6 35.1 105.5




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023



#91  
1,2-Dichlorobenzene  
Concen: 3.086 ug/l  
RT: 13.839 min Scan# 2083  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39


Tgt Ion:146 Resp: 4085  
Ion Ratio Lower Upper  
146 100  
111 39.1 20.4 61.2  
148 64.0 31.9 95.5





#92  
1,2-Dibromo-3-Chloropropane  
Concen: 1.562 ug/l  
RT: 14.451 min Scan# 2187  
Delta R.T. -0.000 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06




Tgt Ion: 75 Resp: 133  
Ion Ratio Lower Upper  
75 100  
155 48.1 43.4 130.2  
157 43.6 56.4 169.2

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023




#93  
1,2,4-Trichlorobenzene  
Concen: 2.723 ug/l  
RT: 15.104 min Scan# 2298  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Tgt Ion:180 Resp: 2367  
Ion Ratio Lower Upper  
180 100  
182 86.9 46.2 138.6  
145 30.4 15.8 47.4





#94  
Hexachlorobutadiene  
Concen: 2.618 ug/l  
RT: 15.210 min Scan# 2316  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Instrument :  
MSVOA\_D  
ClientSampleId :  
MDL06




Tgt Ion:225 Resp: 111  
Ion Ratio Lower Upper  
225 100  
223 67.0 31.8 95.4  
227 56.4 30.7 92.1

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 01/27/2023  
Supervised By :Mahesh Dadoda 01/27/2023




#95  
Naphthalene  
Concen: 2.849 ug/l  
RT: 15.339 min Scan# 2338  
Delta R.T. 0.006 min  
Lab File: VD075212.D  
Acq: 26 Jan 2023 15:39




Tgt Ion:128 Resp: 4930  
Ion Ratio Lower Upper  
128 100  
127 14.0 10.6 15.8  
129 9.5 8.7 13.1





#96  
 1,2,3-Trichlorobenzene  
 Concen: 2.802 ug/l  
 RT: 15.528 min Scan# 21  
 Delta R.T. 0.006 min  
 Lab File: VD075212.D  
 Acq: 26 Jan 2023 15:39

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 MDL06



| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 180     | 100   |       |       |
| 182     | 94.4  | 47.0  | 141.0 |
| 145     | 37.0  | 16.1  | 48.3  |

Manual Integrations  
 APPROVED

Reviewed By :Krupa Patel 01/27/2023  
 Supervised By :Mahesh Dadoda 01/27/2023

