

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD012720\
 Data File : VD064965.D
 Acq On : 27 Jan 2020 15:56
 Operator : VA/SY
 Sample : L1242-06
 Misc : 5.05G/5.00ml/MSVOA D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20200054-FIELD-DUP-COMP

Quant Time: Jan 27 16:31:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D012320S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 23 13:45:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	439754	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	714678	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	638039	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	248169	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	224407	53.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.40%	
35) Dibromofluoromethane	7.92	113	226786	50.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.00%	
50) Toluene-d8	10.34	98	839628	48.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.88%	
62) 4-Bromofluorobenzene	12.64	95	259206	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
Target Compounds						
16) Acetone	4.16	43	149101	144.255	ug/l	97
17) Carbon Disulfide	4.40	76	91094	7.231	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD012720\
Data File : VD064965.D
Acq On : 27 Jan 2020 15:56
Operator : VA/SY
Sample : L1242-06
Misc : 5.05G/5.00ml/MSVOA D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20200054-FIELD-DUP-COMP

Quant Time: Jan 27 16:31:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D012320S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 23 13:45:59 2020
Response via : Initial Calibration

