

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD012720\
 Data File : VD064969.D
 Acq On : 27 Jan 2020 17:51
 Operator : VA/SY
 Sample : L1242-13RE
 Misc : 5.10G/5.00ml/MSVOA D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20200066-AMENDED-COMP-4RE

Quant Time: Jan 27 22:52:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D012320S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 23 13:45:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	417403	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	687981	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	642145	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	281603	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	227405	56.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.60%	
35) Dibromofluoromethane	7.92	113	111883	25.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	51.76%	
50) Toluene-d8	10.34	98	824237	49.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.80%	
62) 4-Bromofluorobenzene	12.64	95	268031	52.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.98%	
Target Compounds						
16) Acetone	4.17	43	177585	182.080	ug/l	99
17) Carbon Disulfide	4.40	76	305831	25.575	ug/l	98
25) 2-Butanone	7.23	43	45873	42.409	ug/l	92
95) Naphthalene	15.41	128	24856	2.902	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD012720\
Data File : VD064969.D
Acq On : 27 Jan 2020 17:51
Operator : VA/SY
Sample : L1242-13RE
Misc : 5.10G/5.00ml/MSVOA D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20200066-AMENDED-COMP-4RE

Quant Time: Jan 27 22:52:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D012320S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 23 13:45:59 2020
Response via : Initial Calibration

