

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012722\
 Data File : VD071807.D
 Acq On : 27 Jan 2022 17:45
 Operator : VA/SY
 Sample : N1267-24
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 2ND-3RD-FLOOR-WALLS-2

Quant Time: Jan 28 07:22:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	401570	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	795691	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	739185	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	323294	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	295571	58.233	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.460%
35) Dibromofluoromethane	7.914	113	264566	50.583	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.160%
50) Toluene-d8	10.338	98	1013998	51.911	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.820%
62) 4-Bromofluorobenzene	12.626	95	324502	47.572	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.140%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012722\
Data File : VD071807.D
Acq On : 27 Jan 2022 17:45
Operator : VA/SY
Sample : N1267-24
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
2ND-3RD-FLOOR-WALLS-2

Quant Time: Jan 28 07:22:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.979 min Scan# 1030
Delta R.T. 0.000 min
Lab File: VD071807.D
Acq: 27 Jan 2022 17:45




Instrument :
MSVOA_D
ClientSampleId :
2ND-3RD-FLOOR-WALLS-2


Tgt Ion:168 Resp: 401570
Ion Ratio Lower Upper
168 100
99 63.9 49.8 74.6



#33
1,2-Dichloroethane-d4
Concen: 58.233 ug/l
RT: 8.332 min Scan# 1090
Delta R.T. 0.006 min
Lab File: VD071807.D
Acq: 27 Jan 2022 17:45

Tgt Ion: 65 Resp: 295571
Ion Ratio Lower Upper
65 100
67 51.2 0.0 101.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.861 min Scan# 1180

Delta R.T. 0.000 min

Lab File: VD071807.D

Acq: 27 Jan 2022 17:45

Instrument :

MSVOA_D

ClientSampleId :

2ND-3RD-FLOOR-WALLS-2





Tgt Ion:114 Resp: 795691

Ion Ratio Lower Upper

114 100

63 21.9 0.0 46.6

88 15.3 0.0 33.2



#35

Dibromofluoromethane

Concen: 50.583 ug/l

RT: 7.914 min Scan# 1019

Delta R.T. 0.000 min

Lab File: VD071807.D

Acq: 27 Jan 2022 17:45




Tgt Ion:113 Resp: 264566


Ion Ratio Lower Upper

113 100

111 104.6 83.8 125.6

192 17.6 14.6 21.8









#50
Toluene-d8
Concen: 51.911 ug/l
RT: 10.338 min Scan# 1431
Delta R.T. 0.000 min
Lab File: VD071807.D
Acq: 27 Jan 2022 17:45




Instrument :
MSVOA_D
ClientSampleId :
2ND-3RD-FLOOR-WALLS-2


Tgt Ion: 98 Resp: 1013998
Ion Ratio Lower Upper
98 100
100 65.6 52.4 78.6



#62
4-Bromofluorobenzene
Concen: 47.572 ug/l
RT: 12.626 min Scan# 1820
Delta R.T. 0.000 min
Lab File: VD071807.D
Acq: 27 Jan 2022 17:45

Tgt Ion: 95 Resp: 324502
Ion Ratio Lower Upper
95 100
174 76.2 0.0 152.8
176 72.5 0.0 145.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.638 min Scan# 1170
Delta R.T. 0.000 min
Lab File: VD071807.D
Acq: 27 Jan 2022 17:45

Instrument :
MSVOA_D
ClientSampleId :
2ND-3RD-FLOOR-WALLS-2

Tgt Ion:117 Resp: 739185
Ion Ratio Lower Upper
117 100
82 58.9 45.3 67.9
119 32.5 26.6 40.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.561 min Scan# 1979
Delta R.T. -0.006 min
Lab File: VD071807.D
Acq: 27 Jan 2022 17:45

Tgt Ion:152 Resp: 323294
Ion Ratio Lower Upper
152 100
115 61.1 31.3 93.8
150 156.1 0.0 348.6

