

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012921\
 Data File : VD068242.D
 Acq On : 29 Jan 2021 16:20
 Operator : VA/SY
 Sample : M1286-10
 Misc : 5.28G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 JJ BH 02 0-1

Quant Time: Jan 30 00:03:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	456752	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	795864	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	685351	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.579	152	287497	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	230665	52.79	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 105.58%			
35) Dibromofluoromethane	7.920	113	233059	52.97	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 105.94%			
50) Toluene-d8	10.344	98	973258	55.30	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 110.60%			
62) 4-Bromofluorobenzene	12.632	95	301765	47.61	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 95.22%			
Target Compounds						Qvalue
16) Acetone	4.156	43	5902	7.57	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012921\
Data File : VD068242.D
Acq On : 29 Jan 2021 16:20
Operator : VA/SY
Sample : M1286-10
Misc : 5.28G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
JJ BH 02 0-1

Quant Time: Jan 30 00:03:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

