

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012921\
 Data File : VD068248.D
 Acq On : 29 Jan 2021 19:04
 Operator : VA/SY
 Sample : M1286-22
 Misc : 4.11G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 JJ BH 03_4-6

Quant Time: Jan 30 00:05:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	336785	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	598233	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	542096	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	239561	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	196906	61.12	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 122.24%		
35) Dibromofluoromethane	7.920	113	177919	53.79	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 107.58%		
50) Toluene-d8	10.344	98	742827	56.15	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 112.30%		
62) 4-Bromofluorobenzene	12.632	95	245230	51.47	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 102.94%		
Target Compounds						Qvalue
16) Acetone	4.162	43	6682	11.62	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012921\
Data File : VD068248.D
Acq On : 29 Jan 2021 19:04
Operator : VA/SY
Sample : M1286-22
Misc : 4.11G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
JJ BH 03_4-6

Quant Time: Jan 30 00:05:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

