

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD013122\
 Data File : VD071845.D
 Acq On : 31 Jan 2022 15:11
 Operator : VA/SY
 Sample : N1325-07
 Misc : 3.33G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TES-3

Quant Time: Jan 31 15:41:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	146230	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	268147	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	291075	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	134533	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	171028	92.534	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 185.060%#	
35) Dibromofluoromethane	7.914	113	113471	64.377	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 128.760%	
50) Toluene-d8	10.338	98	349502	53.094	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.180%	
62) 4-Bromofluorobenzene	12.626	95	129335	56.262	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 112.520%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD013122\
Data File : VD071845.D
Acq On : 31 Jan 2022 15:11
Operator : VA/SY
Sample : N1325-07
Misc : 3.33G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TES-3

Quant Time: Jan 31 15:41:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

