

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD013122\
 Data File : VD071851.D
 Acq On : 31 Jan 2022 17:58
 Operator : VA/SY
 Sample : N1335-01
 Misc : 12.76G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MH-3S

Quant Time: Feb 01 04:04:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	439251	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	787054	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	715454	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	298556	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	255935	46.099	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.200%
35) Dibromofluoromethane	7.914	113	250647	48.448	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.900%
50) Toluene-d8	10.338	98	1009248	52.235	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.460%
62) 4-Bromofluorobenzene	12.626	95	311260	46.131	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.260%
Target Compounds						
16) Acetone	4.168	43	14831	9.759	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD013122\
Data File : VD071851.D
Acq On : 31 Jan 2022 17:58
Operator : VA/SY
Sample : N1335-01
Misc : 12.76G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MH-3S

Quant Time: Feb 01 04:04:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

