

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD013122\
 Data File : VD071853.D
 Acq On : 31 Jan 2022 18:54
 Operator : VA/SY
 Sample : N1335-07
 Misc : 9.71G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MH-4S

Quant Time: Feb 01 04:05:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	397220	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	740596	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	690582	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	290417	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	251226	50.038	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 100.080%	
35) Dibromofluoromethane	7.914	113	236604	48.602	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 97.200%	
50) Toluene-d8	10.338	98	950592	52.285	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.580%	
62) 4-Bromofluorobenzene	12.626	95	305253	48.079	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 96.160%	
Target Compounds						
16) Acetone	4.150	43	17164	12.489	ug/l	Qvalue # 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD013122\
Data File : VD071853.D
Acq On : 31 Jan 2022 18:54
Operator : VA/SY
Sample : N1335-07
Misc : 9.71G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MH-4S

Quant Time: Feb 01 04:05:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

