

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD013122\
 Data File : VD071857.D
 Acq On : 31 Jan 2022 20:45
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Feb 01 04:06:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	444339	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	817009	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	767213	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	334513	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	270858	48.228	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.460%
35) Dibromofluoromethane	7.914	113	255425	47.561	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.120%
50) Toluene-d8	10.338	98	1042789	51.992	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.980%
62) 4-Bromofluorobenzene	12.626	95	344400	49.171	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD013122\
Data File : VD071857.D
Acq On : 31 Jan 2022 20:45
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Feb 01 04:06:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

