

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD020119\
 Data File : VD060828.D
 Acq On : 1 Feb 2019 19:26
 Operator : VA/AP
 Sample : K1312-01RE
 Misc : 6.24g/5ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SOIL-PILERE

Quant Time: Feb 02 00:26:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D012819S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 29 13:53:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.76	168	2617	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	7.95	114	4490	50.00	ug/l	-0.02
63) Chlorobenzene-d5	12.15	117	4212	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	14.30	152	1582	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.19	65	2216	55.23	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	110.46%	
35) Dibromofluoromethane	6.64	113	1055	24.50	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	49.00%	
50) Toluene-d8	10.16	98	3463	34.44	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	68.88%	
62) 4-Bromofluorobenzene	13.34	95	1910	37.79	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	75.58%	
Target Compounds						
					Qvalue	
16) Acetone	3.09	43	422	72.824	ug/l #	59
59) 2-Hexanone	11.27	43	205	13.489	ug/l #	27
67) Ethyl Benzene	12.43	91	297	1.716	ug/l #	49
81) trans-1,4-Dichloro-2-buten	13.34	75	1334	176.381	ug/l #	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD020119\
Data File : VD060828.D
Acq On : 1 Feb 2019 19:26
Operator : VA/AP
Sample : K1312-01RE
Misc : 6.24g/5ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SOIL-PILERE

Quant Time: Feb 02 00:26:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D012819S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 29 13:53:55 2019
Response via : Initial Calibration

