

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020122\
 Data File : VD071866.D
 Acq On : 01 Feb 2022 12:23
 Operator : VA/SY
 Sample : N1340-02
 Misc : 6.88G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-2(1-1.5)

Quant Time: Feb 02 07:05:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	476955	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	868586	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	786410	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	331398	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	281409	46.680	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.360%
35) Dibromofluoromethane	7.914	113	275918	48.326	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.660%
50) Toluene-d8	10.338	98	1100426	51.608	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.220%
62) 4-Bromofluorobenzene	12.626	95	345841	46.445	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.900%
Target Compounds						
16) Acetone	4.162	43	43655	26.454	ug/l	97
25) 2-Butanone	7.220	43	11871	6.171	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020122\
Data File : VD071866.D
Acq On : 01 Feb 2022 12:23
Operator : VA/SY
Sample : N1340-02
Misc : 6.88G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-2(1-1.5)

Quant Time: Feb 02 07:05:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

