

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020122\
 Data File : VD071869.D
 Acq On : 01 Feb 2022 13:46
 Operator : VA/SY
 Sample : N1340-06
 Misc : 7.24G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-6(1.5-2)

Quant Time: Feb 02 07:06:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	447079	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	824293	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	739957	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	279106	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	270569	47.881	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.760%
35) Dibromofluoromethane	7.914	113	264072	48.737	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.480%
50) Toluene-d8	10.338	98	1053375	52.056	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.120%
62) 4-Bromofluorobenzene	12.626	95	306271	43.341	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.680%
Target Compounds						
17) Carbon Disulfide	4.420	76	42924	2.506	ug/l	96
39) Methylcyclohexane	9.355	83	17758	1.648	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020122\
Data File : VD071869.D
Acq On : 01 Feb 2022 13:46
Operator : VA/SY
Sample : N1340-06
Misc : 7.24G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-6(1.5-2)

Quant Time: Feb 02 07:06:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

