

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020321\
 Data File : VD068254.D
 Acq On : 03 Feb 2021 11:15
 Operator : VA/SY
 Sample : VD0203SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0203SBL01

Quant Time: Feb 03 23:39:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	522237	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	911437	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.650	117	793152	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.579	152	346313	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	262804	52.60	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.20%
35) Dibromofluoromethane	7.914	113	260351	51.67	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.34%
50) Toluene-d8	10.344	98	1111375	55.14	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.28%
62) 4-Bromofluorobenzene	12.632	95	354351	48.82	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.64%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020321\
Data File : VD068254.D
Acq On : 03 Feb 2021 11:15
Operator : VA/SY
Sample : VD0203SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0203SBL01

Quant Time: Feb 03 23:39:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

