

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020321\
 Data File : VD068257.D
 Acq On : 03 Feb 2021 12:45
 Operator : VA/SY
 Sample : M1292-01
 Misc : 8.51G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OU421-005-0510

Quant Time: Feb 03 23:40:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	475425	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	831211	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	718279	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.579	152	277067	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	259565	57.07	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 114.14%			
35) Dibromofluoromethane	7.920	113	248244	54.02	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 108.04%			
50) Toluene-d8	10.343	98	1014423	55.19	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 110.38%			
62) 4-Bromofluorobenzene	12.632	95	305811	46.20	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 92.40%			
Target Compounds						
16) Acetone	4.161	43	12733	15.68	ug/l	98
17) Carbon Disulfide	4.414	76	32526	1.94	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020321\
Data File : VD068257.D
Acq On : 03 Feb 2021 12:45
Operator : VA/SY
Sample : M1292-01
Misc : 8.51G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OU421-005-0510

Quant Time: Feb 03 23:40:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

