

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020321\
 Data File : VD068264.D
 Acq On : 03 Feb 2021 15:55
 Operator : VA/SY
 Sample : M1292-08
 Misc : 5.14G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OU421-003-1015

Quant Time: Feb 03 23:42:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	441549	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	790298	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.650	117	708733	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.579	152	311716	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	241430	57.16	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.32%
35) Dibromofluoromethane	7.920	113	239336	54.78	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.56%
50) Toluene-d8	10.344	98	991031	56.71	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	113.42%
62) 4-Bromofluorobenzene	12.632	95	312881	49.71	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.42%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020321\
Data File : VD068264.D
Acq On : 03 Feb 2021 15:55
Operator : VA/SY
Sample : M1292-08
Misc : 5.14G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OU421-003-1015

Quant Time: Feb 03 23:42:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

