

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020321\
 Data File : VD068267.D
 Acq On : 03 Feb 2021 17:17
 Operator : VA/SY
 Sample : M1292-11
 Misc : 5.42G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OU421-009-0510

Quant Time: Feb 03 23:43:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	415686	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	767579	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	685914	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.579	152	301346	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	234753	59.03	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.06%
35) Dibromofluoromethane	7.914	113	230444	54.30	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.60%
50) Toluene-d8	10.344	98	952050	56.09	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.18%
62) 4-Bromofluorobenzene	12.632	95	309062	50.56	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.12%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020321\
Data File : VD068267.D
Acq On : 03 Feb 2021 17:17
Operator : VA/SY
Sample : M1292-11
Misc : 5.42G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OU421-009-0510

Quant Time: Feb 03 23:43:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

