

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020321\
 Data File : VD068270.D
 Acq On : 03 Feb 2021 18:39
 Operator : VA/SY
 Sample : M1292-14
 Misc : 5.82G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OU421-010-0205

Quant Time: Feb 03 23:44:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	404865	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	759160	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	684349	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.579	152	297782	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	237552	61.34	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.68%
35) Dibromofluoromethane	7.920	113	226102	53.87	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.74%
50) Toluene-d8	10.344	98	947248	56.42	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.84%
62) 4-Bromofluorobenzene	12.632	95	297704	49.24	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.48%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020321\
Data File : VD068270.D
Acq On : 03 Feb 2021 18:39
Operator : VA/SY
Sample : M1292-14
Misc : 5.82G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OU421-010-0205

Quant Time: Feb 03 23:44:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

