



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020321\
Data File : VD068252.D
Acq On : 03 Feb 2021 09:19
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Title : SW846 8260
Last Update : Tue Jan 19 13:19:08 2021



AutoFind: Scans 1820, 1821, 1822; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.3	20813	PASS
75	95	30	60	50.6	57587	PASS
95	95	100	100	100.0	113715	PASS
96	95	5	9	6.9	7890	PASS
173	174	0.00	2	1.0	887	PASS
174	95	50	100	75.2	85539	PASS
175	174	5	9	7.4	6350	PASS
176	174	95	101	97.5	83360	PASS
177	176	5	9	7.4	6143	PASS