

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020322\
 Data File : VD071890.D
 Acq On : 03 Feb 2022 13:52
 Operator : VA/SY
 Sample : N1382-04
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WASTE-BUILDING

Quant Time: Feb 04 00:50:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	231479	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	432322	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	437833	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	195320	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	200598	68.562	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	137.120%
35) Dibromofluoromethane	7.909	113	161192	56.722	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.440%
50) Toluene-d8	10.332	98	556945	52.477	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.960%
62) 4-Bromofluorobenzene	12.620	95	191659	51.713	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020322\
Data File : VD071890.D
Acq On : 03 Feb 2022 13:52
Operator : VA/SY
Sample : N1382-04
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WASTE-BUILDING

Quant Time: Feb 04 00:50:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

