

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020325\
 Data File : VD080225.D
 Acq On : 03 Feb 2025 13:10
 Operator : RP/MD
 Sample : VIBLK
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Feb 04 02:08:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D020325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 04 02:06:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	138663	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	246408	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	203836	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	83291	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	91964	50.985	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.980%
35) Dibromofluoromethane	7.810	113	84767	48.539	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.080%
50) Toluene-d8	10.269	98	255643	39.474	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	78.940%
62) 4-Bromofluorobenzene	12.575	95	70663	34.006	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	68.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020325\
Data File : VD080225.D
Acq On : 03 Feb 2025 13:10
Operator : RP/MD
Sample : VIBLK
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Feb 04 02:08:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D020325S.M
Quant Title : SW846 8260
QLast Update : Tue Feb 04 02:06:21 2025
Response via : Initial Calibration

