

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020325\
 Data File : VD080230.D
 Acq On : 03 Feb 2025 17:57
 Operator : RP/MD
 Sample : Q1271-01
 Misc : 5.00G/5ml/MSVOA_D/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RBR200030

Quant Time: Feb 04 06:10:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D020325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 04 05:44:14 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	53607	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	127589	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	131288	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.521	152	51348	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	58242	83.523	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	167.040%	#
35) Dibromofluoromethane	7.810	113	47593	52.632	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	105.260%	
50) Toluene-d8	10.269	98	128533	38.329	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	76.660%	
62) 4-Bromofluorobenzene	12.574	95	45527	42.312	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	84.620%	
Target Compounds						
16) Acetone	4.034	43	2179m	19.389	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020325\
Data File : VD080230.D
Acq On : 03 Feb 2025 17:57
Operator : RP/MD
Sample : Q1271-01
Misc : 5.00G/5ml/MSVOA_D/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RBR200030

Quant Time: Feb 04 06:10:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D020325S.M
Quant Title : SW846 8260
QLast Update : Tue Feb 04 05:44:14 2025
Response via : Initial Calibration

