

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020421\
 Data File : VD068281.D
 Acq On : 04 Feb 2021 17:19
 Operator : VA/SY
 Sample : M1291-21
 Misc : 5.96G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OU421-005-0002

Quant Time: Feb 05 00:05:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	474086	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	886108	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	797634	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	334979	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	273518	60.31	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.62%
35) Dibromofluoromethane	7.914	113	265291	54.15	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.30%
50) Toluene-d8	10.343	98	1111215	56.71	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	113.42%
62) 4-Bromofluorobenzene	12.631	95	345077	48.90	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020421\
Data File : VD068281.D
Acq On : 04 Feb 2021 17:19
Operator : VA/SY
Sample : M1291-21
Misc : 5.96G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OU421-005-0002

Quant Time: Feb 05 00:05:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

