

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020421\
 Data File : VD068290.D
 Acq On : 04 Feb 2021 21:24
 Operator : VA/SY
 Sample : M1315-02
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-1A

Quant Time: Feb 05 00:08:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	459770	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	855759	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	767638	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	338762	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	249420	56.71	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.42%
35) Dibromofluoromethane	7.920	113	251151	53.08	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.16%
50) Toluene-d8	10.343	98	1064760	56.26	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.52%
62) 4-Bromofluorobenzene	12.632	95	335937	49.29	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.58%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020421\
Data File : VD068290.D
Acq On : 04 Feb 2021 21:24
Operator : VA/SY
Sample : M1315-02
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-1A

Quant Time: Feb 05 00:08:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

