

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020421\
 Data File : VD068284.D
 Acq On : 04 Feb 2021 18:41
 Operator : VA/SY
 Sample : M1291-02
 Misc : 10.28G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OU421-007-0205

Quant Time: Feb 05 00:05:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.979	168	409913	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	776567	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	700331	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	287060	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	258316	65.88	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.76%
35) Dibromofluoromethane	7.920	113	239148	55.70	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.40%
50) Toluene-d8	10.344	98	980627	57.10	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.20%
62) 4-Bromofluorobenzene	12.632	95	309514	50.05	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.10%
Target Compounds						
16) Acetone	4.156	43	14453	20.64	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020421\
Data File : VD068284.D
Acq On : 04 Feb 2021 18:41
Operator : VA/SY
Sample : M1291-02
Misc : 10.28G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OU421-007-0205

Quant Time: Feb 05 00:05:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.979 min Scan# 1030
Delta R.T. -0.006 min
Lab File: VD068284.D
Acq: 04 Feb 2021 18:41




Instrument :
MSVOA_D
ClientSampleId :
OU421-007-0205


Tgt Ion:168 Resp: 409913
Ion Ratio Lower Upper
168 100
99 59.1 43.5 65.3



#16
Acetone
Concen: 20.64 ug/l
RT: 4.156 min Scan# 380
Delta R.T. -0.006 min
Lab File: VD068284.D
Acq: 04 Feb 2021 18:41

Tgt Ion: 43 Resp: 14453
Ion Ratio Lower Upper
43 100
58 31.8 24.2 36.4









#33
1,2-Dichloroethane-d4
Concen: 65.88 ug/l
RT: 8.332 min Scan# 1090
Delta R.T. -0.006 min
Lab File: VD068284.D
Acq: 04 Feb 2021 18:41




Instrument :
MSVOA_D
ClientSampleId :
OU421-007-0205


Tgt Ion: 65 Resp: 258316
Ion Ratio Lower Upper
65 100
67 52.7 0.0 105.0



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.867 min Scan# 1181
Delta R.T. 0.000 min
Lab File: VD068284.D
Acq: 04 Feb 2021 18:41

Tgt Ion: 114 Resp: 776567
Ion Ratio Lower Upper
114 100
63 19.8 0.0 34.8
88 15.3 0.0 31.6









#35
Dibromofluoromethane
Concen: 55.70 ug/l
RT: 7.920 min Scan# 1020
Delta R.T. 0.000 min
Lab File: VD068284.D
Acq: 04 Feb 2021 18:41




Instrument :
MSVOA_D
ClientSampleId :
OU421-007-0205


Tgt Ion:113 Resp: 239148
Ion Ratio Lower Upper
113 100
111 103.5 81.4 122.0
192 17.8 16.1 24.1



#50
Toluene-d8
Concen: 57.10 ug/l
RT: 10.344 min Scan# 1432
Delta R.T. 0.000 min
Lab File: VD068284.D
Acq: 04 Feb 2021 18:41

Tgt Ion: 98 Resp: 980627
Ion Ratio Lower Upper
98 100
100 64.2 53.0 79.4









#62
4-Bromofluorobenzene
Concen: 50.05 ug/l
RT: 12.632 min Scan# 1821
Delta R.T. 0.000 min
Lab File: VD068284.D
Acq: 04 Feb 2021 18:41




Instrument :
MSVOA_D
ClientSampleId :
OU421-007-0205


Tgt Ion: 95 Resp: 309514
Ion Ratio Lower Upper
95 100
174 76.0 0.0 172.4
176 73.8 0.0 168.4



#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.644 min Scan# 1653
Delta R.T. -0.005 min
Lab File: VD068284.D
Acq: 04 Feb 2021 18:41


Tgt Ion: 117 Resp: 700331
Ion Ratio Lower Upper
117 100
82 55.9 41.4 62.0
119 31.3 26.1 39.1





#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.573 min Scan# 1981
Delta R.T. 0.000 min
Lab File: VD068284.D
Acq: 04 Feb 2021 18:41

Instrument :
MSVOA_D
ClientSampleId :
OU421-007-0205



Tgt Ion:152 Resp: 287060
Ion Ratio Lower Upper
152 100
115 59.6 28.9 86.8
150 158.7 0.0 344.4

