

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020422\
 Data File : VD071900.D
 Acq On : 04 Feb 2022 15:55
 Operator : VA/SY
 Sample : N1397-02
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 KTS1-MH-02-3-4

Quant Time: Feb 07 04:20:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	9327	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.861	114	15836	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	13669	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	4766	50.000	ug/l	-0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	9529	80.831	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	161.660%
35) Dibromofluoromethane	7.909	113	5450	52.356	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.720%
50) Toluene-d8	10.338	98	18780	48.308	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.620%
62) 4-Bromofluorobenzene	12.620	95	5320	39.187	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020422\
Data File : VD071900.D
Acq On : 04 Feb 2022 15:55
Operator : VA/SY
Sample : N1397-02
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
KTS1-MH-02-3-4

Quant Time: Feb 07 04:20:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

