

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020422\
 Data File : VD071901.D
 Acq On : 04 Feb 2022 16:24
 Operator : VA/SY
 Sample : N1401-01
 Misc : 4.31G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RO-124035

Quant Time: Feb 07 04:20:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	412453	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	764091	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	714286	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	314812	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	267461	51.305	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.600%
35) Dibromofluoromethane	7.908	113	250307	49.836	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.680%
50) Toluene-d8	10.338	98	966641	51.533	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.060%
62) 4-Bromofluorobenzene	12.626	95	306350	46.768	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020422\
Data File : VD071901.D
Acq On : 04 Feb 2022 16:24
Operator : VA/SY
Sample : N1401-01
Misc : 4.31G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RO-124035

Quant Time: Feb 07 04:20:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

