

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020422\
 Data File : VD071902.D
 Acq On : 04 Feb 2022 16:52
 Operator : VA/SY
 Sample : N1401-03
 Misc : 3.68G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PSC-124055

Quant Time: Feb 07 04:20:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	103048	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	181428	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	191850	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	88386	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	104907	80.544	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 161.080%	
35) Dibromofluoromethane	7.909	113	75574	63.370	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 126.740%	
50) Toluene-d8	10.332	98	234684	52.692	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.380%	
62) 4-Bromofluorobenzene	12.626	95	83321	53.570	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 107.140%	
Target Compounds						
16) Acetone	4.173	43	3533	9.909	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071902\
Data File : VD071902.D
Acq On : 04 Feb 2022 16:52
Operator : VA/SY
Sample : N1401-03
Misc : 3.68G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PSC-124055

Quant Time: Feb 07 04:20:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

