

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020422\
 Data File : VD071905.D
 Acq On : 04 Feb 2022 18:18
 Operator : VA/SY
 Sample : N1397-02RE
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 KTS1-MH-02-3-4RE

Quant Time: Feb 07 04:21:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	167177	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	304894	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	299200	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	116646	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	156073	73.862	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 147.720%	
35) Dibromofluoromethane	7.914	113	117824	58.790	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 117.580%	
50) Toluene-d8	10.332	98	384502	51.371	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 102.740%	
62) 4-Bromofluorobenzene	12.626	95	128642	49.216	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 98.440%	
Target Compounds						
					Qvalue	
16) Acetone	4.156	43	58999	102.000	ug/l	90
25) 2-Butanone	7.214	43	11450	16.980	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020422\
Data File : VD071905.D
Acq On : 04 Feb 2022 18:18
Operator : VA/SY
Sample : N1397-02RE
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
KTS1-MH-02-3-4RE

Quant Time: Feb 07 04:21:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

