


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020422\
 Data File : VD071894.D
 Acq On : 04 Feb 2022 10:48
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Title : SW846 8260
 Last Update : Thu Jan 27 06:04:38 2022



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1809

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.5	61520	PASS
75	95	30	60	54.8	156885	PASS
95	95	100	100	100.0	286421	PASS
96	95	5	9	6.2	17758	PASS
173	174	0.00	2	1.6	3538	PASS
174	95	50	100	75.5	216149	PASS
175	174	5	9	7.7	16540	PASS
176	174	95	101	95.0	205248	PASS
177	176	5	9	6.4	13115	PASS