



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020425\
Data File : VD080232.D
Acq On : 04 Feb 2025 09:25
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D020325S.M
Title : SW846 8260
Last Update : Tue Feb 04 05:44:14 2025



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	6756	PASS
75	95	30	60	50.9	17917	PASS
95	95	100	100	100.0	35211	PASS
96	95	5	9	7.0	2454	PASS
173	174	0.00	2	1.5	462	PASS
174	95	50	100	88.6	31211	PASS
175	174	5	9	8.1	2514	PASS
176	174	95	101	99.6	31083	PASS
177	176	5	9	6.1	1909	PASS