

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD020520\
 Data File : VD065070.D
 Acq On : 05 Feb 2020 18:10
 Operator : VA/SY
 Sample : L1408-07
 Misc : 5.49G/5.00ml/MSVOA D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SHORE-RD-PATH-BH-03-A

Quant Time: Feb 06 06:07:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D020520S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 05 15:53:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	58611	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	87261	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.66	117	89417	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	43177	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	35952	72.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	144.34%	
35) Dibromofluoromethane	7.92	113	30252	62.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	124.28%	
50) Toluene-d8	10.34	98	104997	54.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.36%	
62) 4-Bromofluorobenzene	12.64	95	37375	60.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.74%	
Target Compounds						
68) m/p-Xylenes	11.86	106	2431	2.027	ug/l	89
78) n-propylbenzene	12.83	91	3651	1.052	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD020520\
Data File : VD065070.D
Acq On : 05 Feb 2020 18:10
Operator : VA/SY
Sample : L1408-07
Misc : 5.49G/5.00ml/MSVOA D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SHORE-RD-PATH-BH-03-A

Quant Time: Feb 06 06:07:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D020520S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 05 15:53:00 2020
Response via : Initial Calibration

