

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020524\
 Data File : VD078350.D
 Acq On : 05 Feb 2024 17:05
 Operator : RP/MD
 Sample : P1329-15
 Misc : 5.12G/10ml/MSVOA_D/SOIL/B
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 C0FR6

Quant Time: Feb 06 00:53:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM020124SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Feb 06 00:51:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	158796	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	157144	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	63212	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.269	65	77442	17.189	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	68.760%
7) Chloroethane-d5	2.799	69	71152	18.788	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.160%
11) 1,1-Dichloroethene-d2	3.911	65	15114	15.537	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.160%
21) 2-Butanone-d5	6.981	46	23751	35.806	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	71.620%
24) Chloroform-d	7.563	84	83617	18.468	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	73.880%
26) 1,2-Dichloroethane-d4	8.228	65	46467	19.018	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	76.080%
32) Benzene-d6	8.199	84	160704	17.490	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	69.960%
36) 1,2-Dichloropropane-d6	9.210	67	55502	19.620	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	78.480%
41) Toluene-d8	10.269	98	133767	16.208	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	64.840%
43) trans-1,3-Dichloroprop...	10.522	79	23468	18.580	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	74.320%
47) 2-Hexanone-d5	10.875	63	20001	34.849	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	69.700%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	48693	18.524	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	74.080%
66) 1,2-Dichlorobenzene-d4	13.816	152	47456	20.970	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	83.880%
Target Compounds						Qvalue
16) Methylene chloride	4.799	84	15083	4.843	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020524\
Data File : VD078350.D
Acq On : 05 Feb 2024 17:05
Operator : RP/MD
Sample : P1329-15
Misc : 5.12G/10ml/MSVOA_D/SOIL/B
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
C0FR6

Quant Time: Feb 06 00:53:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM020124SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Feb 06 00:51:05 2024
Response via : Initial Calibration

