

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020524\
 Data File : VD078351.D
 Acq On : 05 Feb 2024 17:32
 Operator : RP/MD
 Sample : P1329-14RE
 Misc : 4.75G/10ml/MSVOA_D/SOIL/B
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 C0FR5RE

Quant Time: Feb 06 00:53:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM020124SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Feb 06 00:51:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	153874	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	145990	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	49392	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.269	65	77416	17.733	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.920%
7) Chloroethane-d5	2.799	69	69145	18.842	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.360%
11) 1,1-Dichloroethene-d2	3.910	65	15791	16.752	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.000%
21) 2-Butanone-d5	6.987	46	24049	37.415	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.820%
24) Chloroform-d	7.569	84	81849	18.656	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	74.640%
26) 1,2-Dichloroethane-d4	8.228	65	46547	19.660	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	78.640%
32) Benzene-d6	8.198	84	156702	18.357	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	73.440%
36) 1,2-Dichloropropane-d6	9.210	67	53582	20.389	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	81.560%
41) Toluene-d8	10.269	98	128695	16.785	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	67.120%
43) trans-1,3-Dichloroprop...	10.522	79	22375	19.068	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.280%
47) 2-Hexanone-d5	10.875	63	20140	37.772	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.540%
56) 1,1,2,2-Tetrachloroeth...	12.645	84	49735	20.366	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.480%
66) 1,2-Dichlorobenzene-d4	13.816	152	36969	20.907	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	83.640%
Target Compounds						
16) Methylene chloride	4.810	84	12778	4.234	ug/L	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020524\
Data File : VD078351.D
Acq On : 05 Feb 2024 17:32
Operator : RP/MD
Sample : P1329-14RE
Misc : 4.75G/10ml/MSVOA_D/SOIL/B
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
C0FR5RE

Quant Time: Feb 06 00:53:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM020124SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Feb 06 00:51:05 2024
Response via : Initial Calibration

