


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD020620\
Data File : VD065077.D
Acq On : 06 Feb 2020 10:19
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D020520S.M
Title : SW846 8260
Last Update : Wed Feb 05 15:53:00 2020



AutoFind: Scans 1822, 1823, 1824; Background Corrected with Scan 1813

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.6	42928	PASS
75	95	30	60	53.0	110344	PASS
95	95	100	100	100.0	208384	PASS
96	95	5	9	6.6	13819	PASS
173	174	0.00	2	1.5	2800	PASS
174	95	50	100	92.1	191914	PASS
175	174	5	9	7.5	14379	PASS
176	174	95	101	97.7	187413	PASS
177	176	5	9	6.8	12807	PASS