



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080254\
Data File : VD080254.D
Acq On : 07 Feb 2025 09:20
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D020325S.M
Title : SW846 8260
Last Update : Tue Feb 04 05:44:14 2025



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1858

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.3	8611	PASS
75	95	30	60	51.0	24029	PASS
95	95	100	100	100.0	47117	PASS
96	95	5	9	6.5	3067	PASS
173	174	0.00	2	1.1	475	PASS
174	95	50	100	89.6	42235	PASS
175	174	5	9	7.4	3134	PASS
176	174	95	101	97.8	41299	PASS
177	176	5	9	6.5	2675	PASS