

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020821\
 Data File : VD068316.D
 Acq On : 08 Feb 2021 13:51
 Operator : VA/SY
 Sample : M1332-01
 Misc : 11.48G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-05-020521

Quant Time: Feb 09 00:14:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	275	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.867	114	169	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	66	50.00	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.00	ug/l	-13.57

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0	0.00	ug/l	
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.00%#
35) Dibromofluoromethane	0.000	113	0	0.00	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.00%#
50) Toluene-d8	10.344	98	229	61.27	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	122.54%
62) 4-Bromofluorobenzene	0.000	95	0	0.00	ug/l	
Spiked Amount	50.000	Range	25 - 144	Recovery	=	0.00%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020821\
Data File : VD068316.D
Acq On : 08 Feb 2021 13:51
Operator : VA/SY
Sample : M1332-01
Misc : 11.48G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-05-020521

Quant Time: Feb 09 00:14:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

