

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020821\
 Data File : VD068317.D
 Acq On : 08 Feb 2021 14:42
 Operator : VA/SY
 Sample : M1332-01
 Misc : 6.76G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-05-020521

Quant Time: Feb 09 00:15:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	13348	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	22595	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	18557	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	7086	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	8510	66.65	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	133.30%
35) Dibromofluoromethane	7.914	113	5481	43.88	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.76%
50) Toluene-d8	10.344	98	24680	49.39	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.78%
62) 4-Bromofluorobenzene	12.632	95	7379	41.01	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	82.02%
Target Compounds					Qvalue	
52) Toluene	10.408	92	1717	4.14	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020821\
Data File : VD068317.D
Acq On : 08 Feb 2021 14:42
Operator : VA/SY
Sample : M1332-01
Misc : 6.76G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-05-020521

Quant Time: Feb 09 00:15:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

