

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020821\
 Data File : VD068319.D
 Acq On : 08 Feb 2021 15:51
 Operator : VA/SY
 Sample : M1339-06
 Misc : 5.58G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-5-10

Quant Time: Feb 09 00:15:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	549856	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	971015	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	878998	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	392862	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	296731	56.41	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.82%
35) Dibromofluoromethane	7.920	113	296011	55.14	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.28%
50) Toluene-d8	10.344	98	1200666	55.91	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.82%
62) 4-Bromofluorobenzene	12.632	95	396046	51.22	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.44%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020821\
Data File : VD068319.D
Acq On : 08 Feb 2021 15:51
Operator : VA/SY
Sample : M1339-06
Misc : 5.58G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-5-10

Quant Time: Feb 09 00:15:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

