

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020821\
 Data File : VD068322.D
 Acq On : 08 Feb 2021 17:13
 Operator : VA/SY
 Sample : M1339-01
 Misc : 5.73G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-0-5

Quant Time: Feb 09 00:16:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	517683	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	909881	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	804009	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	338776	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	273050	55.14	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 110.28%	
35) Dibromofluoromethane	7.920	113	272550	54.18	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.36%	
50) Toluene-d8	10.344	98	1125566	55.94	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 111.88%	
62) 4-Bromofluorobenzene	12.632	95	349592	48.25	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 96.50%	
Target Compounds						
16) Acetone	4.162	43	11778	13.32	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020821\
Data File : VD068322.D
Acq On : 08 Feb 2021 17:13
Operator : VA/SY
Sample : M1339-01
Misc : 5.73G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-0-5

Quant Time: Feb 09 00:16:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

