



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020821\
 Data File : VD068310.D
 Acq On : 08 Feb 2021 09:04
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Title : SW846 8260
 Last Update : Tue Jan 19 13:19:08 2021



AutoFind: Scans 1820, 1821, 1822; Background Corrected with Scan 1810

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.5	34589	PASS
75	95	30	60	49.8	92952	PASS
95	95	100	100	100.0	186581	PASS
96	95	5	9	6.8	12766	PASS
173	174	0.00	2	0.9	1381	PASS
174	95	50	100	80.1	149461	PASS
175	174	5	9	7.6	11301	PASS
176	174	95	101	96.7	144533	PASS
177	176	5	9	6.8	9861	PASS