

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021021\
 Data File : VD068346.D
 Acq On : 10 Feb 2021 15:01
 Operator : VA/SY
 Sample : M1363-01
 Misc : 3.39G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VCP-SS-7-SOIL

Quant Time: Feb 11 00:13:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	455761	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	835084	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	726901	50.00	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.579	152	541914	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	257121	58.97	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 117.94%	
35) Dibromofluoromethane	7.920	113	254513	55.13	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.26%	
50) Toluene-d8	10.344	98	1415236	76.63	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 153.26%#	
62) 4-Bromofluorobenzene	12.632	95	955347	143.65	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 287.30%#	
Target Compounds						
					Qvalue	
16) Acetone	4.162	43	33326	42.81	ug/l	95
39) Methylcyclohexane	9.355	83	443081	40.24	ug/l	90
73) Isopropylbenzene	12.485	105	4074164	90.35	ug/l	100
78) n-propylbenzene	12.820	91	9075496	170.12	ug/l	98
85) sec-Butylbenzene	13.402	105	10947763	245.94	ug/l	89
89) n-Butylbenzene	13.844	91	7046488	181.36	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021021\
Data File : VD068346.D
Acq On : 10 Feb 2021 15:01
Operator : VA/SY
Sample : M1363-01
Misc : 3.39G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VCP-SS-7-SOIL

Quant Time: Feb 11 00:13:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

