

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021021\
 Data File : VD068348.D
 Acq On : 10 Feb 2021 16:49
 Operator : VA/SY
 Sample : M1364-04
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BP-TT-VPB-RW6-GW-378-380

Quant Time: Feb 11 00:14:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	531534	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	937072	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	802121	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	281302	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	267764	52.66	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 105.32%		
35) Dibromofluoromethane	7.920	113	279570	53.96	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 107.92%		
50) Toluene-d8	10.343	98	1129287	54.50	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 109.00%		
62) 4-Bromofluorobenzene	12.632	95	324714	43.51	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 87.02%		
Target Compounds						Qvalue
16) Acetone	4.161	43	8468	9.33	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021021\
Data File : VD068348.D
Acq On : 10 Feb 2021 16:49
Operator : VA/SY
Sample : M1364-04
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BP-TT-VPB-RW6-GW-378-380

Quant Time: Feb 11 00:14:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

