

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021021\
 Data File : VD068339.D
 Acq On : 10 Feb 2021 11:30
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 11 00:09:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	65	0.00
2 T	Dichlorodifluoromethane	0.594	0.502	15.5	56	0.00
3 P	Chloromethane	0.640	0.595	7.0	63	0.00
4 C	Vinyl Chloride	0.711	0.734	-3.2#	67	0.00
5 T	Bromomethane	0.488	0.480	1.6	65	0.00
6 T	Chloroethane	0.435	0.463	-6.4	68	0.00
7 T	Trichlorofluoromethane	0.934	0.909	2.7	63	0.00
8 T	Diethyl Ether	0.275	0.252	8.4	59	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.545	0.549	-0.7	65	0.00
10 T	Methyl Iodide	0.527	0.545	-3.4	57	0.00
11 T	Tert butyl alcohol	0.065	0.025	61.5#	38#	0.00
12 CM	1,1-Dichloroethene	0.548	0.524	4.4#	61	0.00
13 T	Acrolein	0.038	0.031	18.4	57	0.00
14 T	Allyl chloride	0.797	0.708	11.2	56	0.00
15 T	Acrylonitrile	0.109	0.104	4.6	59	0.00
16 T	Acetone	0.085	0.097	-14.1	71	0.00
17 T	Carbon Disulfide	1.764	1.537	12.9	56	0.00
18 T	Methyl Acetate	0.222	0.224	-0.9	64	0.00
19 T	Methyl tert-butyl Ether	1.266	1.125	11.1	56	0.00
20 T	Methylene Chloride	0.684	0.598	12.6	63	0.00
21 T	trans-1,2-Dichloroethene	0.629	0.605	3.8	62	0.00
22 T	Diisopropyl ether	1.608	1.523	5.3	60	0.00
23 T	Vinyl Acetate	0.857	0.782	8.8	56	0.00
24 P	1,1-Dichloroethane	1.042	1.042	0.0	64	0.00
25 T	2-Butanone	0.129	0.117	9.3	58	0.00
26 T	2,2-Dichloropropane	0.986	0.907	8.0	59	0.00
27 T	cis-1,2-Dichloroethene	0.683	0.673	1.5	63	0.00
28 T	Bromochloromethane	0.355	0.330	7.0	65	0.00
29 T	Tetrahydrofuran	0.083	0.074	10.8	56	0.00
30 C	Chloroform	1.053	1.049	0.4#	64	0.00
31 T	Cyclohexane	1.036	0.929	10.3	60	0.00
32 T	1,1,1-Trichloroethane	0.949	0.946	0.3	64	0.00
33 S	1,2-Dichloroethane-d4	0.478	0.401	16.1	58	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	65	0.00
35 S	Dibromofluoromethane	0.276	0.253	8.3	61	0.00
36 T	1,1-Dichloropropene	0.504	0.504	0.0	65	0.00
37 T	Ethyl Acetate	0.178	0.160	10.1	57	0.00
38 T	Carbon Tetrachloride	0.472	0.482	-2.1	65	0.00
39 T	Methylcyclohexane	0.659	0.633	3.9	61	0.00
40 TM	Benzene	1.410	1.407	0.2	63	0.00
41 T	Methacrylonitrile	0.108	0.091	15.7	51	0.00
42 TM	1,2-Dichloroethane	0.377	0.362	4.0	61	0.00
43 T	Isopropyl Acetate	0.372	0.303	18.5	51	0.00
44 TM	Trichloroethene	0.387	0.379	2.1	62	0.00
45 C	1,2-Dichloropropane	0.345	0.341	1.2#	63	0.00
46 T	Dibromomethane	0.175	0.173	1.1	61	0.00
47 T	Bromodichloromethane	0.472	0.472	0.0	62	0.00
48 T	Methyl methacrylate	0.185	0.158	14.6	52	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021021\
 Data File : VD068339.D
 Acq On : 10 Feb 2021 11:30
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 11 00:09:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	57	0.00
50 S	Toluene-d8	1.106	0.917	17.1	59	0.00
51 T	4-Methyl-2-Pentanone	0.173	0.156	9.8	56	0.00
52 CM	Toluene	0.919	0.902	1.8#	63	0.00
53 T	t-1,3-Dichloropropene	0.478	0.454	5.0	59	0.00
54 T	cis-1,3-Dichloropropene	0.569	0.538	5.4	60	0.00
55 T	1,1,2-Trichloroethane	0.248	0.248	0.0	63	0.00
56 T	Ethyl methacrylate	0.330	0.295	10.6	56	0.00
57 T	1,3-Dichloropropane	0.437	0.419	4.1	60	0.00
58 T	2-Chloroethyl Vinyl ether	0.152	0.153	-0.7	68	0.00
59 T	2-Hexanone	0.120	0.113	5.8	59	0.00
60 T	Dibromochloromethane	0.311	0.300	3.5	61	0.00
61 T	1,2-Dibromoethane	0.245	0.232	5.3	60	0.00
62 S	4-Bromofluorobenzene	0.398	0.331	16.8	57	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	64	0.00
64 T	Tetrachloroethene	0.352	0.337	4.3	62	0.00
65 PM	Chlorobenzene	1.062	1.058	0.4	63	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.364	2.7	60	0.00
67 C	Ethyl Benzene	1.967	1.961	0.3#	63	0.00
68 T	m/p-Xylenes	0.743	0.733	1.3	63	0.00
69 T	o-Xylene	0.690	0.681	1.3	63	0.00
70 T	Styrene	1.177	1.149	2.4	61	0.00
71 P	Bromoform	0.189	0.179	5.3	59	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	65	0.00
73 T	Isopropylbenzene	4.160	4.161	-0.0	64	0.00
74 T	N-amyl acetate	0.865	0.715	17.3	53	0.00
75 P	1,1,2,2-Tetrachloroethane	0.662	0.650	1.8	62	0.00
76 T	1,2,3-Trichloropropane	0.483	0.456	5.6	55	0.00
77 T	Bromobenzene	0.892	0.855	4.1	61	0.00
78 T	n-propylbenzene	4.922	4.932	-0.2	64	0.00
79 T	2-Chlorotoluene	2.764	2.669	3.4	62	0.00
80 T	1,3,5-Trimethylbenzene	3.465	3.375	2.6	63	0.00
81 T	trans-1,4-Dichloro-2-butene	0.239	0.206	13.8	55	0.00
82 T	4-Chlorotoluene	2.874	2.790	2.9	62	0.00
83 T	tert-Butylbenzene	2.936	2.936	0.0	63	0.00
84 T	1,2,4-Trimethylbenzene	3.439	3.307	3.8	62	0.00
85 T	sec-Butylbenzene	4.107	4.173	-1.6	65	0.00
86 T	p-Isopropyltoluene	3.736	3.736	0.0	64	0.00
87 T	1,3-Dichlorobenzene	1.715	1.705	0.6	64	0.00
88 T	1,4-Dichlorobenzene	1.690	1.647	2.5	62	0.00
89 T	n-Butylbenzene	3.585	3.695	-3.1	66	0.00
90 T	Hexachloroethane	0.675	0.678	-0.4	63	0.00
91 T	1,2-Dichlorobenzene	1.469	1.446	1.6	62	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.108	0.095	12.0	56	0.00
93 T	1,2,4-Trichlorobenzene	1.007	0.965	4.2	60	0.00
94 T	Hexachlorobutadiene	0.558	0.548	1.8	63	0.00
95 T	Naphthalene	1.888	1.763	6.6	59	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021021\
Data File : VD068339.D
Acq On : 10 Feb 2021 11:30
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Feb 11 00:09:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.849	0.811	4.5	60	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6